

The National Marine Biological  
Analytical Quality Control Scheme

[www.nmbaqcs.org](http://www.nmbaqcs.org)

Particle Size Results – PS46

**Authors:** Adam Procter & Sarah Hussey

Reviewed by: Richard Arnold

Approved by: Richard Arnold

Contact: Adam Procter

[adam.procter@unicomarine.com](mailto:adam.procter@unicomarine.com)

Thomson Unicomarine Ltd.

Date of Issue: 28th May 2013

## Contents

### **Tables**

- Table 1. Summary of the replicate benchmark analysis and particle size information received from participating laboratories for the forty-sixth PSA NMBAQC Scheme.
- Table 2. Summary of z-scores for each phi-interval for PS46; data from all participating laboratories included in the mean and standard deviation calculations.

### **Figures**

- Figure 1. Benchmark particle size distribution curves for PS46 resulting from analysis of ten replicate samples.
- Figure 2. Particle size distribution curves from all participating laboratories for sediment samples from PS46.
- Figure 3. Summary of z-scores for the benchmark data (TUM AVERAGE); when data from all participating laboratories are included in mean and standard deviation calculations.
- Figure 4. Cluster dendrogram of PS46 including all laboratories, with the benchmark replicates (TUM average).
- Figure 5. MDS plots of PS46 with the benchmark replicates (TUM AVERAGE) averaged; (a) including all laboratories and (b) an exploded subset of cluster group b.

### **Appendices**

- Appendix 1. Final Summary Data sheets as supplied by participating laboratories (arranged by Lab Code).
- Appendix 2. Z-score calculations when data from all participating laboratories are included in mean and standard deviation calculations.
- Appendix 3. Summary of z-scores for each half-phi interval for PS46; when data from all participating laboratories included in the mean and standard deviation calculations.

Table 1. Summary of the replicate benchmark analysis and particle size information received from participating laboratories for the forty-sixth PSA NMBAQC Scheme

Benchmark Data

Sample	Method	% Gravel	% Sand	% Silt	Median $\phi$	Mean $\phi$	Sediment Description (Post analysis)
PS46 1960	NMBAQC	98.76	1.23	0.01	-2.839	-2.988	Gravel
PS46 1961	NMBAQC	98.75	1.24	0.01	-2.844	-2.995	Gravel
PS46 1962	NMBAQC	98.24	1.72	0.04	-2.829	-2.981	Gravel
PS46 1963	NMBAQC	98.98	1.01	0.00	-2.871	-3.015	Gravel
PS46 1964	NMBAQC	99.08	0.91	0.02	-2.844	-2.986	Gravel
PS46 1965	NMBAQC	98.97	1.01	0.01	-2.834	-2.987	Gravel
PS46 1966	NMBAQC	98.57	1.39	0.03	-2.838	-2.983	Gravel
PS46 1967	NMBAQC	98.72	1.27	0.00	-2.876	-3.002	Gravel
PS46 1968	NMBAQC	98.96	1.00	0.03	-2.825	-2.962	Gravel
PS46 1969	NMBAQC	98.32	1.65	0.03	-2.833	-2.979	Gravel
TUM AVERAGE	NMBAQC	98.74	1.24	0.02	-2.843	-2.988	

Participant Data

Lab	Method	% Gravel	% Sand	% Silt	Sediment Description (Post analysis)
LB_1901	NMBAQC	98.49	1.49	0.01	Gravel
LB_1903	NMBAQC	98.77	1.21	0.02	Gravel
LB_1904	NMBAQC	99.15	0.85	0.00	Gravel
LB_1905	NMBAQC	98.83	1.07	0.10	Gravel
LB_1908	OTHER	99.32	0.67	0.01	Gravel
LB_1909	NMBAQC	99.05	0.95	0.00	Gravel
LB_1910	NMBAQC	98.21	1.75	0.05	Gravel
LB_1917	NMBAQC	99.22	0.78	0.00	Gravel
LB_1921	NMBAQC	91.71	8.20	0.09	Gravel
LB_1955	NMBAQC	99.47	0.52	0.01	Gravel
LB_1958	NMBAQC	99.21	0.79	0.00	Gravel

Key to methods

NMBAQC - States following NMBAQC PSA SOP for supporting biological data

OTHER - Following a different SOP.

Figure 1. Benchmark particle size distribution curves for PS46 resulting from analysis of ten replicate samples.




Figure 2. Particle size distribution curves from all participating laboratories for sediment samples from PS46.




Table 2. Summary of z-scores for each half-phi interval for PS46; data from all participating laboratories included in mean and standard deviation calculations.


	-6.50 to -6.00	-6.00 to -5.50	-5.50 to -5.00	-5.00 to -4.50	-4.50 to -4.00	-4.00 to -3.50	-3.50 to -3.00	-3.00 to -2.50	-2.50 to -2.00	-2.00 to -1.50	-1.50 to -1.00	-1.00 to -0.50	-0.50 to 0.00	0.00 to 0.50	0.50 to 1.00	1.00 to 1.50	1.50 to 2.00	2.00 to 2.50	2.50 to 3.00	3.00 to 3.50
TUM AVERAGE	0.000	-0.315	0.000	-0.373	0.598893	0.392314	-0.450226	-1.331689	1.198029	0.871975	-0.769393	-0.201382	1.135524	-0.325688	-0.439492	-0.508044	-0.507679	-0.535717	-0.545724	-0.597589
LB1901	0.000	-0.314918	0.000	-0.372641	0.799915	0.340257	-0.436898	0.054926	0.188728	0.147047	-1.787937	-0.106717	<b>2.512321</b>	0.158774	-0.220649	-0.504915	-0.57171	-0.632834	-0.815591	-0.698844
LB1903	0.000	<b>3.149183</b>	0.000	-0.372641	-0.367095	0.483377	-0.517807	-0.757784	1.110628	-0.028397	0.001898	-0.233429	0.192709	1.266875	<b>2.268801</b>	<b>2.163058</b>	1.209647	1.374654	1.075355	1.721417
LB1904	0.000	-0.314918	0.000	-0.372641	<b>-2.548926</b>	<b>-3.03089</b>	<b>3.07502</b>	1.148197	0.114402	-0.519517	0.663015	-0.340667	0.121438	-0.805094	-0.832812	-0.809588	-0.708663	-0.786599	-0.761839	-0.930601
LB1905	0.000	-0.314918	0.000	-0.372641	0.508833	0.570018	-0.529356	-1.57973	0.86863	<b>2.01546</b>	0.388347	-0.27133	0.043232	0.395744	0.333511	0.135379	-0.053193	0.158186	0.317032	1.253715
LB1908	0.000	-0.314918	0.000	-0.372641	0.063674	0.673623	-0.256638	0.460107	-0.331452	0.569621	1.374739	-0.439912	-1.041054	-0.556156	-0.649501	-0.617149	-0.354243	-0.431645	-0.466597	-0.662168
LB1909	0.000	-0.314918	0.000	-0.372641	0.803462	-0.039152	-0.222366	-1.329057	1.264883	-0.820474	0.737569	-0.278914	-0.648801	-0.805094	-0.832812	-0.809588	-0.708663	-0.786599	-0.761839	-0.930601
LB1910	0.000	-0.314918	0.000	-0.372641	0.568453	0.440322	-0.49878	0.677453	-0.173078	0.748674	-0.131246	-0.221532	0.60572	0.394407	0.75153	1.560687	<b>2.806465</b>	<b>2.363499</b>	<b>2.55462</b>	1.059299
LB1917	0.000	-0.314918	0.000	-0.372641	0.816925	0.377042	-0.201178	0.611752	-1.044293	-0.448429	0.534083	-0.356718	-0.998801	-0.805094	-0.832812	-0.809588	-0.708663	-0.786599	-0.761839	-0.930601
LB1921	0.000	-0.314918	0.000	0.262836	-0.112831	0.181074	-0.348716	1.289732	<b>-2.054537</b>	<b>-0.963557</b>	-1.891573	<b>3.138594</b>	0.215558	-0.775341	-0.411518	-0.101004	-0.169658	-0.092811	0.023815	0.7148
LB1955	0.000	-0.314918	0.000	-0.372641	0.555919	-0.294028	0.181684	-0.701859	0.575877	-1.578144	0.475542	-0.526436	-0.632709	<b>2.338074</b>	1.259075	0.602295	0.159343	0.407149	0.158721	0.335385
LB1958	0.000	-0.314918	0.000	<b>3.090934</b>	-1.088328	0.298157	-0.244965	0.126265	-0.517787	0.877717	-0.362236	-0.362939	-0.369813	-0.805094	-0.832812	-0.809588	-0.708663	-0.786599	-0.761839	-0.930601
Mean	0.000	0.342732	0.000	0.96227	12.08068	19.20398	9.283497	30.93658	16.80542	3.476638	5.05593	1.491105	0.066897	0.019852	0.022723	0.024687	0.030087	0.026702	0.028745	0.02128
St. Dev	0.000	1.088319	0.000	2.582296	4.739518	8.336466	10.25129	1.820241	2.60916	0.510274	0.264083	2.088788	0.032621	0.024658	0.027285	0.030716	0.042576	0.033946	0.037731	0.022867

	3.50 to 4.00	4.00 to 4.50	4.50 to 5.00	5.00 to 5.50	5.50 to 6.00	6.00 to 6.50	6.50 to 7.00	7.00 to 7.50	7.50 to 8.00	8.00 to 8.50	8.50 to 9.00	9.00 to 9.50	9.50 to 10.00	10.00 to 10.50	10.50 to 11.00	11.00 to 11.50	11.50 to 12.00	12.00 to 12.50	12.50 to 13.00	13.00 to 13.50
TUM AVERAGE	-0.65624	-0.613233	-0.410296	-0.410121	-0.42125	-0.434128	-0.443665	-0.449925	-0.452197	-0.454816	-0.455803	-0.45776	-0.438003	-0.404905	-0.365653	-0.362117	-0.314918	-0.314918	-0.314918	-0.314918
LB1901	-0.709639	-0.672006	-0.478428	-0.464597	-0.464018	-0.471383	-0.478459	-0.475967	-0.471728	-0.467255	-0.456696	-0.43769	-0.438003	-0.404905	-0.365653	-0.362117	-0.314918	-0.314918	-0.314918	-0.314918
LB1903	0.88286	0.450801	0.227478	-0.012664	-0.1105	-0.175471	-0.231689	-0.276464	-0.296902	-0.29808	-0.301838	-0.29981	-0.259711	-0.186852	-0.127067	-0.362117	-0.314918	-0.314918	-0.314918	-0.314918
LB1904	-0.884216	-0.827258	-0.607361	-0.562535	-0.541298	-0.535054	-0.532101	-0.525279	-0.517969	-0.510554	-0.494685	-0.468027	-0.438003	-0.404905	-0.365653	-0.362117	-0.314918	-0.314918	-0.314918	-0.314918
LB1905	1.374347	<b>1.998253</b>	<b>2.63738</b>	<b>2.668167</b>	<b>2.578771</b>	<b>2.3981</b>	<b>2.20114</b>	1.956379	1.759109	1.592372	1.318874	0.91283	0.541902	0.138546	-0.303214	-0.362117	-0.314918	-0.314918	-0.314918	-0.314918
LB1908	-0.639419	-0.653956	-0.453647	-0.477311	-0.506126	-0.517102	-0.512319	-0.503131	-0.496854	-0.493453	-0.481966	-0.455834	-0.426437	-0.393583	-0.362939	-0.362117	-0.314918	-0.314918	-0.314918	-0.314918
LB1909	-0.884216	-0.827258	-0.607361	-0.562535	-0.541298	-0.535054	-0.532101	-0.525279	-0.517969	-0.510554	-0.494685	-0.468027	-0.438003	-0.404905	-0.365653	-0.362117	-0.314918	-0.314918	-0.314918	-0.314918
LB1910	1.837818	1.326582	-0.607361	-0.562535	-0.541298	-0.535054	-0.532101	-0.525279	-0.517969	-0.510554	-0.494685	-0.468027	-0.438003	-0.404905	-0.365653	-0.362117	-0.314918	-0.314918	-0.314918	-0.314918
LB1917	-0.884216	-0.827258	-0.607361	-0.562535	-0.541298	-0.535054	-0.532101	-0.525279	-0.517969	-0.510554	-0.494685	-0.468027	-0.438003	-0.404905	-0.365653	-0.362117	-0.314918	-0.314918	-0.314918	-0.314918
LB1921	0.717035	1.026885	1.241072	1.290029	1.472858	1.741889	<b>1.981145</b>	<b>2.226956</b>	<b>2.393862</b>	<b>2.516422</b>	<b>2.688661</b>	<b>2.88518</b>	<b>3.012075</b>	<b>3.099208</b>	<b>3.132764</b>	<b>3.110898</b>	-0.314918	-0.314918	-0.314918	-0.314918
LB1955	0.07386	-0.167527	-0.137052	-0.190953	-0.264293	-0.300742	-0.301312	-0.298831	-0.29768	-0.297234	-0.293259	-0.27452	-0.239811	-0.227883	-0.145628	0.148157	<b>3.149183</b>	<b>3.149183</b>	<b>3.149183</b>	<b>3.149183</b>
LB1958	-0.884216	-0.827258	-0.607361	-0.562535	-0.541298	-0.535054	-0.532101	-0.525279	-0.517969	-0.510554	-0.494685	-0.468027	-0.438003	-0.404905	-0.365653	-0.362117	-0.314918	-0.314918	-0.314918	-0.314918
Mean	0.022172	0.017477	0.011113	0.009783	0.008835	0.008209	0.007703	0.007391	0.006894	0.006031	0.004939	0.0037	0.002496	0.001969	0.001127	0.000383	3.9E-05	2.97E-05	2.23E-05	1.43E-05
St. Dev	0.025075	0.021127	0.018296	0.017391	0.016322	0.015343	0.014476	0.014057	0.01331	0.011812	0.009985	0.00794	0.005703	0.004862	0.003082	0.001058	0.000124	9.42E-05	7.07E-05	4.55E-05

**z-score >1.96 or <-1.96**  
 All values equal zero

Figure 3. Summary of z-scores for the benchmark data (TUM Average); when data from all participating laboratories are included in mean and standard deviation calculations.



## Results of SIMPROF testing on PSA Ring test PS46 data

Data was entered into PRIMER v. 6.1.13 in half-phi intervals; any missing data was entered as zero. The data did not need to be transformed as all data was on a similar percentage scale. A Euclidean distance matrix was created from the data; The Euclidean distance between two samples (labs)  $j$  and  $k$ , is defined algebraically as  $d_{jk} = \sqrt{\sum_{i=1}^p (y_{ij} - y_{ik})^2}$ . From this distance matrix cluster analysis was carried out including a SIMPROF test at a 5% significance level. The red SIMPROF lines on the dendrogram indicate labs that cannot be distinguished from each other at the 5% significance level; the black lines indicate labs that can be distinguished from each other. The results are presented as a cluster dendrogram (Figure 4) and non-metric Multi-Dimensional Scaling (MDS) diagrams (Figures 5) below. It is important to note that, although the MDS plot is bounded by a box, the box does not represent either axes or scale. Two samples with a high similarity index will appear close together while those less similar will appear further apart. The 'correct' configuration of sample points will be multidimensional and the plot represents the best 2-dimensional solution to the problem. The technique should be viewed as complementary to cluster analysis, offering a different perspective of the same information.

Figure 4. Cluster dendrogram of PS46 including all laboratories, with the benchmark replicates (TUM average).








Figure 5. MDS plots of PS46 with the benchmark replicates (TUM AVERAGE) averaged; (a) including all laboratories and (b) a subset of cluster group b.

5a.



5b.



The cluster analysis separates the laboratories into 2 SIMPROF cluster groups; one of these groups comprises of a single lab.

Cluster group A is formed of the single laboratory (LB1904), figure 2 shows that their cumulative percentage is displaced by one phi and rises sharply between -3.5 and -3 phi. This could be due to a data entry error.

Cluster group B consists of all other laboratories including the TUM average (LB1901, LB1903, LB1905, LB1908, LB1909, LB1910, LB1917, LB1921, LB1955, and LB1958). These laboratories cumulative frequency curves (figure 2) are all fairly similar, with small variations below -4.5 phi for labs LB1903 and LB1958.

## Appendices

Appendix 1. Final Summary Data sheets as supplied by participating laboratories (arranged by Lab Code).

Exercise Code:	PS46
LabCode:	LB1901
Sample Code:	PS461901
Equipment used (e.g. laser model and range):	Endecotts Test Sieves, Malvern Mastersizer 2000 Laser Diffractor (Model: MAL1002178)
Method used:	NMBAQC PSA SOP for supporting biological data*
Peroxide pre-treatment used:	NO*
Chemical dispersant used:	NO*
Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume % (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	15.8719
-4.00 to -3.50; 11.2 mm	21.3599
-3.50 to -3.00; 8 mm	4.8037
-3.00 to -2.50; 5.6 mm	31.0256
-2.50 to -2.00; 4 mm	17.2978
-2.00 to -1.50; 2.8 mm	3.5517
-1.50 to -1.00; 2 mm	4.5838
-1.00 to -0.50; 1.4 mm	1.2682
-0.50 to 0.00; 1 mm	0.1487
0.00 to 0.50; (707 µm)	0.0238
0.50 to 1.00; (500 µm)	0.0167
1.00 to 1.50; (353.6 µm)	0.0094
1.50 to 2.00; (250 µm)	0.0057
2.00 to 2.50; (176.8 µm)	0.0052
2.50 to 3.00; (125 µm)	0.0055
3.00 to 3.50; (88.39 µm)	0.0053
3.50 to 4.00; (62.5 µm)	0.0044
4.00 to 4.50; (44.19 µm)	0.0033
4.50 to 5.00; (31.25 µm)	0.0024
5.00 to 5.50; (22.097 µm)	0.0017
5.50 to 6.00; (15.625 µm)	0.0013
6.00 to 6.50; (11.049 µm)	0.0010
6.50 to 7.00; (7.813 µm)	0.0008
7.00 to 7.50; (5.524 µm)	0.0007
7.50 to 8.00; (3.906 µm)	0.0006
8.00 to 8.50; (2.762 µm)	0.0005
8.50 to 9.00; (1.953 µm)	0.0004
9.00 to 9.50; (1.381 µm)	0.0002
9.50 to 10.00; (0.977 µm)	0.0000
10.00 to 10.50; (0.691 µm)	0.0000
10.50 to 11.00; (0.488 µm)	0.0000
11.00 to 11.50; (0.345 µm)	0.0000
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000

<b>Exercise Code:</b>	<b>PS46</b>
<b>LabCode:</b>	<b>LB1903</b>
<b>Sample Code:</b>	<b>PS461903</b>
<b>Equipment used (e.g. laser model and range):</b>	<i>Malvern 2000 (0.02 - 2000 µm) Hydro G</i>
<b>Method used:</b>	<b>NMBAQC PSA SOP for supporting biological data*</b>
<b>Peroxide pre-treatment used:</b>	<b>NO*</b>
<b>Chemical dispersant used:</b>	<b>NO*</b>
<b>Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)</b>	<b>Volume/Weight (mark as "0" for not analysed or no material)</b>
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	16.5300
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	45.3400
-4.00 to -3.50; 11.2 mm	97.6300
-3.50 to -3.00; 8 mm	17.4300
-3.00 to -2.50; 5.6 mm	130.2600
-2.50 to -2.00; 4 mm	86.3900
-2.00 to -1.50; 2.8 mm	15.1800
-1.50 to -1.00; 2 mm	22.1700
-1.00 to -0.50; 1.4 mm	4.4000
-0.50 to 0.00; 1 mm	0.3200
0.00 to 0.50; (707 µm)	0.0511
0.50 to 1.00; (500 µm)	0.0846
1.00 to 1.50; (353.6 µm)	0.0913
1.50 to 2.00; (250 µm)	0.0816
2.00 to 2.50; (176.8 µm)	0.0734
2.50 to 3.00; (125 µm)	0.0693
3.00 to 3.50; (88.39 µm)	0.0606
3.50 to 4.00; (62.5 µm)	0.0443
4.00 to 4.50; (44.19 µm)	0.0270
4.50 to 5.00; (31.25 µm)	0.0153
5.00 to 5.50; (22.097 µm)	0.0096
5.50 to 6.00; (15.625 µm)	0.0070
6.00 to 6.50; (11.049 µm)	0.0055
6.50 to 7.00; (7.813 µm)	0.0043
7.00 to 7.50; (5.524 µm)	0.0035
7.50 to 8.00; (3.906 µm)	0.0029
8.00 to 8.50; (2.762 µm)	0.0025
8.50 to 9.00; (1.953 µm)	0.0019
9.00 to 9.50; (1.381 µm)	0.0013
9.50 to 10.00; (0.977 µm)	0.0010
10.00 to 10.50; (0.691 µm)	0.0011
10.50 to 11.00; (0.488 µm)	0.0007
11.00 to 11.50; (0.345 µm)	0.0000
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000

Exercise Code:	PS46
LabCode:	LB1904
Sample Code:	PS461904
Equipment used (e.g. laser model and range):	Fritsch Sieve Shaker
Method used:	NMBAQC PSA SOP for supporting biological data*
Peroxide pre-treatment used:	NO*
Chemical dispersant used:	NO*
Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	0.0000
-4.00 to -3.50; 11.2 mm	0.0000
-3.50 to -3.00; 8 mm	179.0300
-3.00 to -2.50; 5.6 mm	143.8900
-2.50 to -2.00; 4 mm	75.0400
-2.00 to -1.50; 2.8 mm	14.0900
-1.50 to -1.00; 2 mm	22.9500
-1.00 to -0.50; 1.4 mm	3.4200
-0.50 to 0.00; 1 mm	0.3100
0.00 to 0.50; (707 µm)	0.0000
0.50 to 1.00; (500 µm)	0.0000
1.00 to 1.50; (353.6 µm)	0.0000
1.50 to 2.00; (250 µm)	0.0000
2.00 to 2.50; (176.8 µm)	0.0000
2.50 to 3.00; (125 µm)	0.0000
3.00 to 3.50; (88.39 µm)	0.0000
3.50 to 4.00; (62.5 µm)	0.0000
4.00 to 4.50; (44.19 µm)	0.0000
4.50 to 5.00; (31.25 µm)	0.0000
5.00 to 5.50; (22.097 µm)	0.0000
5.50 to 6.00; (15.625 µm)	0.0000
6.00 to 6.50; (11.049 µm)	0.0000
6.50 to 7.00; (7.813 µm)	0.0000
7.00 to 7.50; (5.524 µm)	0.0000
7.50 to 8.00; (3.906 µm)	0.0000
8.00 to 8.50; (2.762 µm)	0.0000
8.50 to 9.00; (1.953 µm)	0.0000
9.00 to 9.50; (1.381 µm)	0.0000
9.50 to 10.00; (0.977 µm)	0.0000
10.00 to 10.50; (0.691 µm)	0.0000
10.50 to 11.00; (0.488 µm)	0.0000
11.00 to 11.50; (0.345 µm)	0.0000
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000

Exercise Code:	PS46
LabCode:	LB1905
Sample Code:	PS461905
Equipment used (e.g. laser model and range):	Mastersizer 2000, hydro mu accessory unit, sieve stack (1mm-16mm in half phi intervals)
Method used:	NMBAQC PSA SOP for supporting biological data*
Peroxide pre-treatment used:	NO*
Chemical dispersant used:	NO*
Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	63.9848
-4.00 to -3.50; 11.2 mm	100.7336
-3.50 to -3.00; 8 mm	17.0286
-3.00 to -2.50; 5.6 mm	125.2871
-2.50 to -2.00; 4 mm	84.1807
-2.00 to -1.50; 2.8 mm	19.8903
-1.50 to -1.00; 2 mm	22.7728
-1.00 to -0.50; 1.4 mm	4.0811
-0.50 to 0.00; 1 mm	0.3007
0.00 to 0.50; (707 µm)	0.0296
0.50 to 1.00; (500 µm)	0.0318
1.00 to 1.50; (353.6 µm)	0.0290
1.50 to 2.00; (250 µm)	0.0278
2.00 to 2.50; (176.8 µm)	0.0321
2.50 to 3.00; (125 µm)	0.0407
3.00 to 3.50; (88.39 µm)	0.0499
3.50 to 4.00; (62.5 µm)	0.0566
4.00 to 4.50; (44.19 µm)	0.0597
4.50 to 5.00; (31.25 µm)	0.0594
5.00 to 5.50; (22.097 µm)	0.0562
5.50 to 6.00; (15.625 µm)	0.0509
6.00 to 6.50; (11.049 µm)	0.0450
6.50 to 7.00; (7.813 µm)	0.0396
7.00 to 7.50; (5.524 µm)	0.0349
7.50 to 8.00; (3.906 µm)	0.0303
8.00 to 8.50; (2.762 µm)	0.0248
8.50 to 9.00; (1.953 µm)	0.0181
9.00 to 9.50; (1.381 µm)	0.0109
9.50 to 10.00; (0.977 µm)	0.0056
10.00 to 10.50; (0.691 µm)	0.0026
10.50 to 11.00; (0.488 µm)	0.0002
11.00 to 11.50; (0.345 µm)	0.0000
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000



Exercise Code:	PS46
LabCode:	LB1908
Sample Code:	PS461908
Equipment used (e.g. laser model and range):	Endecotts Test Sieves, Malvern Mastersizer 2000 Laser Diffractor (Model: MAL1002178)
Method used:	Whole sample dry sieved down to <63um, and <63um (Pan) fraction subjected to laser diffraction (based on BS1377: 1990 Parts 1-2 and BS13320: 2009).
Peroxide pre-treatment used:	NO*
Chemical dispersant used:	NO*
Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	12.3825
-4.00 to -3.50; 11.2 mm	23.4723
-3.50 to -3.00; 8 mm	6.6536
-3.00 to -2.50; 5.6 mm	31.6821
-2.50 to -2.00; 4 mm	15.9406
-2.00 to -1.50; 2.8 mm	3.7673
-1.50 to -1.00; 2 mm	5.4189
-1.00 to -0.50; 1.4 mm	0.5722
-0.50 to 0.00; 1 mm	0.0327
0.00 to 0.50; (707 µm)	0.0061
0.50 to 1.00; (500 µm)	0.0050
1.00 to 1.50; (353.6 µm)	0.0059
1.50 to 2.00; (250 µm)	0.0150
2.00 to 2.50; (176.8 µm)	0.0120
2.50 to 3.00; (125 µm)	0.0111
3.00 to 3.50; (88.39 µm)	0.0061
3.50 to 4.00; (62.5 µm)	0.0061
4.00 to 4.50; (44.19 µm)	0.0037
4.50 to 5.00; (31.25 µm)	0.0028
5.00 to 5.50; (22.097 µm)	0.0015
5.50 to 6.00; (15.625 µm)	0.0006
6.00 to 6.50; (11.049 µm)	0.0003
6.50 to 7.00; (7.813 µm)	0.0003
7.00 to 7.50; (5.524 µm)	0.0003
7.50 to 8.00; (3.906 µm)	0.0003
8.00 to 8.50; (2.762 µm)	0.0002
8.50 to 9.00; (1.953 µm)	0.0001
9.00 to 9.50; (1.381 µm)	0.0001
9.50 to 10.00; (0.977 µm)	0.0001
10.00 to 10.50; (0.691 µm)	0.0001
10.50 to 11.00; (0.488 µm)	0.0000
11.00 to 11.50; (0.345 µm)	0.0000
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	"0"
12.50 to 13.00; (0.122 µm)	"0"
13.00 to 13.50; (0.086 µm)	"0"

<b>Exercise Code:</b>	<b>PS46</b>
<b>LabCode:</b>	<b>LB1909</b>
<b>Sample Code:</b>	<b>PS461909</b>
<b>Equipment used (e.g. laser model and range):</b>	<i>Malvern Mastersizer 2000 (0.01µm to 2000µm)</i>
<b>Method used:</b>	<b>NMBAQC PSA SOP for supporting biological data*</b>
<b>Peroxide pre-treatment used:</b>	<b>NO*</b>
<b>Chemical dispersant used:</b>	<b>NO*</b>
<b>Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)</b>	<b>Volume/Weight (mark as "0" for not analysed or no material)</b>
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	69.7800
-4.00 to -3.50; 11.2 mm	83.2500
-3.50 to -3.00; 8 mm	30.7600
-3.00 to -2.50; 5.6 mm	126.4100
-2.50 to -2.00; 4 mm	88.3000
-2.00 to -1.50; 2.8 mm	13.4300
-1.50 to -1.00; 2 mm	23.0600
-1.00 to -0.50; 1.4 mm	3.9900
-0.50 to 0.00; 1 mm	0.2000
0.00 to 0.50; (707 µm)	0.0000
0.50 to 1.00; (500 µm)	0.0000
1.00 to 1.50; (353.6 µm)	0.0000
1.50 to 2.00; (250 µm)	0.0000
2.00 to 2.50; (176.8 µm)	0.0000
2.50 to 3.00; (125 µm)	0.0000
3.00 to 3.50; (88.39 µm)	0.0000
3.50 to 4.00; (62.5 µm)	0.0000
4.00 to 4.50; (44.19 µm)	0.0000
4.50 to 5.00; (31.25 µm)	0.0000
5.00 to 5.50; (22.097 µm)	0.0000
5.50 to 6.00; (15.625 µm)	0.0000
6.00 to 6.50; (11.049 µm)	0.0000
6.50 to 7.00; (7.813 µm)	0.0000
7.00 to 7.50; (5.524 µm)	0.0000
7.50 to 8.00; (3.906 µm)	0.0000
8.00 to 8.50; (2.762 µm)	0.0000
8.50 to 9.00; (1.953 µm)	0.0000
9.00 to 9.50; (1.381 µm)	0.0000
9.50 to 10.00; (0.977 µm)	0.0000
10.00 to 10.50; (0.691 µm)	0.0000
10.50 to 11.00; (0.488 µm)	0.0000
11.00 to 11.50; (0.345 µm)	0.0000
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000

Exercise Code:	PS46
LabCode:	LB1910
Sample Code:	PS461910
Equipment used (e.g. laser model and range):	Retsch AS 200 Sieve Shaker
Method used:	NMBAQC PSA SOP for supporting biological data*
Peroxide pre-treatment used:	NO*
Chemical dispersant used:	NO*
Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	14.7749
-4.00 to -3.50; 11.2 mm	21.9939
-3.50 to -3.00; 8 mm	4.1704
-3.00 to -2.50; 5.6 mm	32.0342
-2.50 to -2.00; 4 mm	16.3538
-2.00 to -1.50; 2.8 mm	3.8587
-1.50 to -1.00; 2 mm	5.0213
-1.00 to -0.50; 1.4 mm	1.0284
-0.50 to 0.00; 1 mm	0.0865
0.00 to 0.50; (707 µm)	0.0296
0.50 to 1.00; (500 µm)	0.0432
1.00 to 1.50; (353.6 µm)	0.0728
1.50 to 2.00; (250 µm)	0.1411
2.00 to 2.50; (176.8 µm)	0.1069
2.50 to 3.00; (125 µm)	0.1251
3.00 to 3.50; (88.39 µm)	0.0455
3.50 to 4.00; (62.5 µm)	0.0683
4.00 to 4.50; (44.19 µm)	0.0455
4.50 to 5.00; (31.25 µm)	0.0000
5.00 to 5.50; (22.097 µm)	0.0000
5.50 to 6.00; (15.625 µm)	0.0000
6.00 to 6.50; (11.049 µm)	0.0000
6.50 to 7.00; (7.813 µm)	0.0000
7.00 to 7.50; (5.524 µm)	0.0000
7.50 to 8.00; (3.906 µm)	0.0000
8.00 to 8.50; (2.762 µm)	0.0000
8.50 to 9.00; (1.953 µm)	0.0000
9.00 to 9.50; (1.381 µm)	0.0000
9.50 to 10.00; (0.977 µm)	0.0000
10.00 to 10.50; (0.691 µm)	0.0000
10.50 to 11.00; (0.488 µm)	0.0000
11.00 to 11.50; (0.345 µm)	0.0000
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000

Exercise Code:	PS46
LabCode:	LB1917
Sample Code:	PS461917
Equipment used (e.g. laser model and range):	Mastersizer with Hydro2000G
Method used:	NMBAQC PSA SOP for supporting biological data*
Peroxide pre-treatment used:	NO*
Chemical dispersant used:	NO*
Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	70.1400
-4.00 to -3.50; 11.2 mm	94.9400
-3.50 to -3.00; 8 mm	31.7500
-3.00 to -2.50; 5.6 mm	140.3800
-2.50 to -2.00; 4 mm	61.9100
-2.00 to -1.50; 2.8 mm	14.2800
-1.50 to -1.00; 2 mm	22.8500
-1.00 to -0.50; 1.4 mm	3.2800
-0.50 to 0.00; 1 mm	0.1500
0.00 to 0.50; (707 µm)	0.0000
0.50 to 1.00; (500 µm)	0.0000
1.00 to 1.50; (353.6 µm)	0.0000
1.50 to 2.00; (250 µm)	0.0000
2.00 to 2.50; (176.8 µm)	0.0000
2.50 to 3.00; (125 µm)	0.0000
3.00 to 3.50; (88.39 µm)	0.0000
3.50 to 4.00; (62.5 µm)	0.0000
4.00 to 4.50; (44.19 µm)	0.0000
4.50 to 5.00; (31.25 µm)	0.0000
5.00 to 5.50; (22.097 µm)	0.0000
5.50 to 6.00; (15.625 µm)	0.0000
6.00 to 6.50; (11.049 µm)	0.0000
6.50 to 7.00; (7.813 µm)	0.0000
7.00 to 7.50; (5.524 µm)	0.0000
7.50 to 8.00; (3.906 µm)	0.0000
8.00 to 8.50; (2.762 µm)	0.0000
8.50 to 9.00; (1.953 µm)	0.0000
9.00 to 9.50; (1.381 µm)	0.0000
9.50 to 10.00; (0.977 µm)	0.0000
10.00 to 10.50; (0.691 µm)	0.0000
10.50 to 11.00; (0.488 µm)	0.0000
11.00 to 11.50; (0.345 µm)	0.0000
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000

Exercise Code:	PS46
LabCode:	LB1921
Sample Code:	PS461921
Equipment used (e.g. laser model and range):	Malvern Mastersizer 2000 MU
Method used:	NMBAQC PSA SOP for supporting biological data*
Peroxide pre-treatment used:	NO*
Chemical dispersant used:	NO*
Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	7.7900
-4.50 to -4.00; 16 mm	54.8100
-4.00 to -3.50; 11.2 mm	96.6100
-3.50 to -3.00; 8 mm	27.1000
-3.00 to -2.50; 5.6 mm	156.7800
-2.50 to -2.00; 4 mm	54.3300
-2.00 to -1.50; 2.8 mm	14.1700
-1.50 to -1.00; 2 mm	21.6300
-1.00 to -0.50; 1.4 mm	38.2000
-0.50 to 0.00; 1 mm	0.3500
0.00 to 0.50; (707 µm)	0.0007
0.50 to 1.00; (500 µm)	0.0115
1.00 to 1.50; (353.6 µm)	0.0218
1.50 to 2.00; (250 µm)	0.0229
2.00 to 2.50; (176.8 µm)	0.0236
2.50 to 3.00; (125 µm)	0.0296
3.00 to 3.50; (88.39 µm)	0.0376
3.50 to 4.00; (62.5 µm)	0.0402
4.00 to 4.50; (44.19 µm)	0.0392
4.50 to 5.00; (31.25 µm)	0.0338
5.00 to 5.50; (22.097 µm)	0.0322
5.50 to 6.00; (15.625 µm)	0.0329
6.00 to 6.50; (11.049 µm)	0.0349
6.50 to 7.00; (7.813 µm)	0.0364
7.00 to 7.50; (5.524 µm)	0.0387
7.50 to 8.00; (3.906 µm)	0.0388
8.00 to 8.50; (2.762 µm)	0.0358
8.50 to 9.00; (1.953 µm)	0.0318
9.00 to 9.50; (1.381 µm)	0.0266
9.50 to 10.00; (0.977 µm)	0.0197
10.00 to 10.50; (0.691 µm)	0.0170
10.50 to 11.00; (0.488 µm)	0.0108
11.00 to 11.50; (0.345 µm)	0.0037
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000

Exercise Code:	PS46
LabCode:	LB1955
Sample Code:	PS461955
Equipment used (e.g. laser model and range):	Coulter LS230 with variable speed fluid module
Method used:	NMBAQC PSA SOP for supporting biological data*
Peroxide pre-treatment used:	NO*
Chemical dispersant used:	NO*
Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	71.7000
-4.00 to -3.50; 11.2 mm	77.3000
-3.50 to -3.00; 8 mm	7.9000
-3.00 to -2.50; 5.6 mm	31.1000
-2.50 to -2.00; 4 mm	1.7000
-2.00 to -1.50; 2.8 mm	0.1000
-1.50 to -1.00; 2 mm	0.0000
-1.00 to -0.50; 1.4 mm	0.0000
-0.50 to 0.00; 1 mm	0.0000
0.00 to 0.50; (707 µm)	0.0097
0.50 to 1.00; (500 µm)	20.4824
1.00 to 1.50; (353.6 µm)	140.1686
1.50 to 2.00; (250 µm)	170.8780
2.00 to 2.50; (176.8 µm)	61.8426
2.50 to 3.00; (125 µm)	8.8291
3.00 to 3.50; (88.39 µm)	2.5097
3.50 to 4.00; (62.5 µm)	0.0000
4.00 to 4.50; (44.19 µm)	0.0000
4.50 to 5.00; (31.25 µm)	0.0000
5.00 to 5.50; (22.097 µm)	0.0000
5.50 to 6.00; (15.625 µm)	0.0000
6.00 to 6.50; (11.049 µm)	0.0000
6.50 to 7.00; (7.813 µm)	0.0000
7.00 to 7.50; (5.524 µm)	0.0000
7.50 to 8.00; (3.906 µm)	0.0000
8.00 to 8.50; (2.762 µm)	0.0000
8.50 to 9.00; (1.953 µm)	0.0000
9.00 to 9.50; (1.381 µm)	0.0000
9.50 to 10.00; (0.977 µm)	0.0000
10.00 to 10.50; (0.691 µm)	0.0000
10.50 to 11.00; (0.488 µm)	0.0000
11.00 to 11.50; (0.345 µm)	0.0000
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000

Exercise Code:	PS46
LabCode:	LB1958
Sample Code:	PS461958
Equipment used (e.g. laser model and range):	
Method used:	NMBAQC PSA SOP for supporting biological data*
Peroxide pre-treatment used:	NO*
Chemical dispersant used:	NO*
Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	39.2900
-4.50 to -4.00; 16 mm	30.4100
-4.00 to -3.50; 11.2 mm	92.6600
-3.50 to -3.00; 8 mm	29.7500
-3.00 to -2.50; 5.6 mm	136.8000
-2.50 to -2.00; 4 mm	67.8900
-2.00 to -1.50; 2.8 mm	17.2400
-1.50 to -1.00; 2 mm	21.7900
-1.00 to -0.50; 1.4 mm	3.2200
-0.50 to 0.00; 1 mm	0.2400
0.00 to 0.50; (707 µm)	0.0000
0.50 to 1.00; (500 µm)	0.0000
1.00 to 1.50; (353.6 µm)	0.0000
1.50 to 2.00; (250 µm)	0.0000
2.00 to 2.50; (176.8 µm)	0.0000
2.50 to 3.00; (125 µm)	0.0000
3.00 to 3.50; (88.39 µm)	0.0000
3.50 to 4.00; (62.5 µm)	0.0000
4.00 to 4.50; (44.19 µm)	0.0000
4.50 to 5.00; (31.25 µm)	0.0000
5.00 to 5.50; (22.097 µm)	0.0000
5.50 to 6.00; (15.625 µm)	0.0000
6.00 to 6.50; (11.049 µm)	0.0000
6.50 to 7.00; (7.813 µm)	0.0000
7.00 to 7.50; (5.524 µm)	0.0000
7.50 to 8.00; (3.906 µm)	0.0000
8.00 to 8.50; (2.762 µm)	0.0000
8.50 to 9.00; (1.953 µm)	0.0000
9.00 to 9.50; (1.381 µm)	0.0000
9.50 to 10.00; (0.977 µm)	0.0000
10.00 to 10.50; (0.691 µm)	0.0000
10.50 to 11.00; (0.488 µm)	0.0000
11.00 to 11.50; (0.345 µm)	0.0000
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000









Appendix 2. Z-score calculations when data from all participating laboratories are included in mean and standard deviation calculations.

	6.00 to 6.50	z-score	6.50 to 7.00	z-score	7.00 to 7.50	z-score	7.50 to 8.00	z-score	8.00 to 8.50	z-score	8.50 to 9.00	z-score
TUM AVERAGE	0.001548	-0.434128	0.00128	-0.443665	0.001066	-0.449925	0.000875	-0.452197	0.000658	-0.454816	0.000388	-0.455803
LB1901	0.000977	-0.471383	0.000805	-0.476459	0.000701	-0.475957	0.000615	-0.471728	0.000511	-0.467255	0.000377	-0.45695
LB1903	0.005517	-0.175471	0.004349	-0.231689	0.003505	-0.276464	0.002942	-0.296902	0.00251	-0.29808	0.001928	-0.301838
LB1904	0	-0.535054	0	-0.532101	0	-0.52579	0	-0.517959	0	-0.510554	0	-0.494685
LB1905	0.045003	<b>2.3981</b>	0.039565	<b>2.20114</b>	0.034893	<b>1.958379</b>	0.030309	1.759109	0.024839	1.592372	0.018108	1.318874
LB1908	0.000275	-0.517102	0.000288	-0.512319	0.000319	-0.503131	0.000281	-0.498854	0.000202	-0.493453	0.000127	-0.481968
LB1909	0	-0.535054	0	-0.532101	0	-0.52579	0	-0.517959	0	-0.510554	0	-0.494685
LB1910	0	-0.535054	0	-0.532101	0	-0.52579	0	-0.517959	0	-0.510554	0	-0.494685
LB1917	0	-0.535054	0	-0.532101	0	-0.52579	0	-0.517959	0	-0.510554	0	-0.494685
LB1921	0.034934	1.741869	0.036381	<b>1.981145</b>	0.036696	<b>2.226956</b>	0.038758	<b>2.393852</b>	0.035754	<b>2.516422</b>	0.031784	<b>2.688561</b>
LB1955	0.003595	-0.300742	0.003341	-0.301312	0.00319	-0.298831	0.002932	-0.29768	0.00252	-0.297234	0.002011	-0.293259
LB1958	0	-0.535054	0	-0.532101	0	-0.52579	0	-0.517959	0	-0.510554	0	-0.494685
Mean	0.008209		0.007703		0.007391		0.006894		0.006031		0.004939	
St. Dev	0.015343		0.014476		0.014057		0.01331		0.011812		0.009985	





	9.00 to 9.50	z-score	9.50 to 10.00	z-score	10.00 to 10.50	z-score	10.50 to 11.00	z-score	11.00 to 11.50	z-score	11.50 to 12.00	z-score
TUM AVERAGE	6.56E-05	-0.45776	0	-0.438003	0	-0.404905	0	-0.365653	0	-0.362117	0	-0.314918
LB1901	0.000225	-0.43769	0	-0.438003	0	-0.404905	0	-0.365653	0	-0.362117	0	-0.314918
LB1903	0.00132	-0.29981	0.001017	-0.259711	0.00108	-0.188852	0.000735	-0.127067	0	-0.362117	0	-0.314918
LB1904	0	-0.468027	0	-0.438003	0	-0.404905	0	-0.365653	0	-0.362117	0	-0.314918
LB1905	0.010948	0.91283	0.005589	0.541902	0.002642	0.138546	0.000192	-0.303214	0	-0.362117	0	-0.314918
LB1908	8.09E-05	-0.455834	6.6E-05	-0.426437	5.5E-05	-0.393583	8.36E-06	-0.362939	0	-0.362117	0	-0.314918
LB1909	0	-0.468027	0	-0.438003	0	-0.404905	0	-0.365653	0	-0.362117	0	-0.314918
LB1910	0	-0.468027	0	-0.438003	0	-0.404905	0	-0.365653	0	-0.362117	0	-0.314918
LB1917	0	-0.468027	0	-0.438003	0	-0.404905	0	-0.365653	0	-0.362117	0	-0.314918
LB1921	0.026608	<b>2.88516</b>	0.019677	<b>3.012075</b>	0.017036	<b>3.099206</b>	0.010781	<b>3.132764</b>	0.003674	<b>3.110898</b>	0	-0.314918
LB1955	0.001521	-0.27452	0.00113	-0.239811	0.000881	-0.227883	0.000678	-0.145628	0.00054	0.148157	0.000428	<b>3.149183</b>
LB1958	0	-0.468027	0	-0.438003	0	-0.404905	0	-0.365653	0	-0.362117	0	-0.314918
Mean	0.0037		0.002498		0.001969		0.001127		0.000383		3.9E-05	
St. Dev	0.00794		0.005703		0.004862		0.003082		0.001058		0.000124	

	12.00 to 12.50	z-score	12.50 to 13.00	z-score	13.00 to 13.50	z-score
TUM AVERAGE		0 -0.314918		0 -0.314918		0 -0.314918
LB1901		0 -0.314918		0 -0.314918		0 -0.314918
LB1903		0 -0.314918		0 -0.314918		0 -0.314918
LB1904		0 -0.314918		0 -0.314918		0 -0.314918
LB1905		0 -0.314918		0 -0.314918		0 -0.314918
LB1908		0 -0.314918		0 -0.314918		0 -0.314918
LB1909		0 -0.314918		0 -0.314918		0 -0.314918
LB1910		0 -0.314918		0 -0.314918		0 -0.314918
LB1917		0 -0.314918		0 -0.314918		0 -0.314918
LB1921		0 -0.314918		0 -0.314918		0 -0.314918
LB1955	0.000326	<b>3.149183</b>	0.000245	<b>3.149183</b>	0.000157	<b>3.149183</b>
LB1958		0 -0.314918		0 -0.314918		0 -0.314918
Mean	2.97E-05		2.23E-05		1.43E-05	
St. Dev	9.42E-05		7.07E-05		4.55E-05	

Appendix 3. Summary of z-scores for each half-phi interval for PS46; when data from all participating laboratories included in the mean and standard deviation calculations.



Appendix 3. Summary of z-scores for each half-phi interval for PS46; when data from all participating laboratories included in the mean and standard deviation calculations.



Appendix 3. Summary of z-scores for each half-phi interval for PS46; when data from all participating laboratories included in the mean and standard deviation calculations.

