

**The National Marine Biological Analytical Quality Control Scheme**

[www.nmbaqcs.org](http://www.nmbaqcs.org)

**Particle Size Results – PS55**

Lydia Finbow and David Hall  
March 2015  
Email: [nmbaqc@apemltd.com](mailto:nmbaqc@apemltd.com)



## **Contents**

### **Tables**

- Table 1. Summary of the methodology and sieve metadata information received from participating laboratories and replicate analysis laboratory for the fifty-fifth particle size distribution - PS55.
- Table 2. Summary of the methodology and <1mm; >0.00 phi metadata received from participating laboratories and replicate analysis laboratory for the fifty-fifth particle size distribution - PS55.
- Table 3. Summary of the particle size information received from participating laboratories and replicate analysis for the fifty-fifth particle size distribution – PS55.
- Table 4. Summary of the z-scores for each half phi-interval for PS55.

### **Figures**

- Figure 1. Particle size distribution curves resulting from analysis of five replicate samples of sediment distributed as PS55 (Benchmark Data).
- Figure 2. Particle size distribution curves from all participating laboratories and the benchmark average for sediment sample from PS55.
- Figure 3. PSA Particle size ternary diagrams for PS55, including the Benchmark replicates and all participating laboratories for Gravel, Sand and Mud.
- Figure 4. Bar chart for PS55 showing the percentage gravel, sand and silt/ clay recorded by each participating laboratory and the benchmark average.
- Figure 5. Summary of z-scores at each half phi-interval for the benchmark data.
- Figures 6 - 8. Summary of z-scores at each half phi-interval for participating laboratories.
- Figure 9. Cluster dendrogram from a Euclidean distance matrix for all participating laboratories with the benchmark replicates averaged.

## **Appendices**

- Appendix 1. Final Merged Data sheets as supplied by participating laboratories (arranged by Lab Code) and the benchmark replicates.
- Appendix 2. Z-score calculations when data from all participating laboratories are included in the mean and standard deviation calculations.

## **PARTICLE SIZE EXERCISE DETAILS**

### **Particle Size (PS) #55**

- Type/Contents/Origin/Pre-treatments:
- ~50g dried sand (Swanage) <0.5 mm pre-sieve
  - 800g Gravel (aggregate source) @ various prescribed ½ phi categories
  - 500g wet mud (Barry Island)
  - 200g of water to mix.

Phi category/ sediment type		Approximate weight in grams
Gravel	3.5 to -3.0	415
	-3.0 to -2.5	340
	2.5 to -2.0	40
	-2.0 to -1.5	5
Sand		50 dried
Mud		500 wet

**Circulated: 18/012/2014**

**Completion Date: 06/02/2015**

**Number of Subscribing Laboratories: 15**

**Number of Participating Laboratories: 13**

**Number of Results Received: 14\***

\* One sample from PSA\_2113 was stored frozen on arrival, due to lack of information on storage and to samples arriving during Christmas break. A replacement sample was later provided, which was not frozen. PSA\_2113 is the non-frozen sample and PSA\_2113\_a is the data from the frozen sample.

**Table 1. Summary of the methodology and sieve metadata information received from participating laboratories and replicate analysis laboratory for the fifty-fifth particle size distribution - PS55.**

Benchmark	Method	Sieves used	Phi; sieve mesh		Total Weight (g)
			Weight (g) < 0.00; >1 mm	Weight (g) > 0.00; <1 mm	
REPLICATE 1	NMBAQC	☑	728.96	179.48	908.44
REPLICATE 2	NMBAQC	☑	724.28	185.39	909.67
REPLICATE 3	NMBAQC	☑	749.04	187.55	936.59
REPLICATE 4	NMBAQC	☑	733.71	183.52	917.23
REPLICATE 5	NMBAQC	☑	734.61	181.36	915.97
BM AVERAGE	NMBAQC	☑	734.12	183.46	917.58

Participant	Method	Sieves used	Phi; sieve mesh		Total Weight (g)
			Weight (g) < 0.00; >1 mm	Weight (g) > 0.00; <1 mm	
PSA_2101	NMBAQC	☑	759.97	194.63	954.60
PSA_2102	NMBAQC	☑	724.08	183.08	907.16
PSA_2103	NMBAQC	☑	795.21	199.92	995.13
PSA_2105	NMBAQC	☑	728.96	179.48	908.44
PSA_2106	NMBAQC	☑	81.3415	0.00	81.3415
PSA_2107	NMBAQC	☑	791.70	179.90	971.60
PSA_2108	NMBAQC <sup>1</sup>	☑	792.19	196.62	988.81
PSA_2109	OTHER	☑	794.08	217.13	1011.21
PSA_2110	NMBAQC	☑	790.87	192.05	982.92
PSA_2111	OTHER	☑	800.12	503.8	1303.92
PSA_2112	NMBAQC	☑	778.22	190.54	968.76
PSA_2113	NMBAQC	☑	792.13	197.19	989.32
PSA_2113_a	NMBAQC	☑	791.36	191.59	982.95
PSA_2114	NMBAQC	☑	752.08	188.96	941.04

**Key to methods**

NMBAQC - NMBAQC PSA SOP for supporting biological data

NMBAQC<sup>1</sup> - NMBAQC PSA SOP for supporting biological data - incorporating BS1377: 1990 Parts 1-2 (sieving) and BS13320: 2009 (laser diffraction)

BLUE TEXT = amended after interim report

**Table 2. Summary of the <1mm; >0.00 phi laser metadata received from participating laboratories and replicate analysis laboratory for the fifty-fifth particle size distribution - PS55.**

Benchmark	Laser Used	% Sand				% Silt				% Clay 8 - 13 phi
		Coarse 0 - 1 phi	Medium 1 - 2 phi	Fine 2 - 3 phi	Very Fine 3 - 4 phi	Coarse 4 - 5 phi	Medium 5 - 6 phi	Fine 6 - 7 phi	Very Fine 7 - 8 phi	
REPLICATE 1	<input checked="" type="checkbox"/>	0.00	5.47	8.85	4.19	6.48	11.44	14.66	17.19	31.72
REPLICATE 2	<input checked="" type="checkbox"/>	0.00	5.56	8.65	3.90	6.56	11.51	15.33	17.90	30.60
REPLICATE 3	<input checked="" type="checkbox"/>	0.00	4.83	9.55	4.13	6.36	11.26	14.81	17.52	31.54
REPLICATE 4	<input checked="" type="checkbox"/>	0.00	5.17	9.67	4.68	6.51	11.27	14.93	17.52	30.25
REPLICATE 5	<input checked="" type="checkbox"/>	0.00	4.76	9.96	3.75	6.42	11.59	14.70	17.17	31.65
BM AVERAGE	<input checked="" type="checkbox"/>	0.00	5.16	9.34	4.13	6.46	11.41	14.89	17.46	31.15

Participant	Laser Used	% Sand				% Silt				% Clay 8 - 13 phi
		Coarse 0 - 1 phi	Medium 1 - 2 phi	Fine 2 - 3 phi	Very Fine 3 - 4 phi	Coarse 4 - 5 phi	Medium 5 - 6 phi	Fine 6 - 7 phi	Very Fine 7 - 8 phi	
PSA_2101	<input checked="" type="checkbox"/>	0.76	8.02	10.92	5.91	7.05	10.71	15.00	15.91	25.73
PSA_2102	<input checked="" type="checkbox"/>	1.01	6.22	5.24	2.19	8.29	14.33	20.30	22.91	19.51
PSA_2103	<input checked="" type="checkbox"/>	3.46	24.03	14.51	6.98	13.06	12.53	9.41	7.01	9.04
PSA_2105	<input checked="" type="checkbox"/>	0.00	5.47	8.85	4.19	6.48	11.44	14.66	17.19	31.72
PSA_2106 *	<input checked="" type="checkbox"/>	1.49	2.48	0.92	0.92	1.93	2.69	2.83	2.19	3.21
PSA_2107	<input checked="" type="checkbox"/>	0.01	2.74	3.76	4.21	7.82	12.68	20.34	23.72	24.72
PSA_2108	<input checked="" type="checkbox"/>	2.06	10.82	8.16	4.40	8.07	12.27	16.42	17.85	19.95
PSA_2109	<input checked="" type="checkbox"/>	-	-	-	-	-	-	-	-	-
PSA_2110	<input checked="" type="checkbox"/>	0.54	4.96	4.60	4.13	7.58	13.44	20.89	23.00	20.86
PSA_2111	<input checked="" type="checkbox"/>	1.51	6.10	4.59	3.10	6.98	12.43	18.79	21.48	25.02
PSA_2112	<input checked="" type="checkbox"/>	0.09	2.90	3.62	3.79	7.42	12.97	21.32	28.24	19.67
PSA_2113	<input checked="" type="checkbox"/>	0.00	4.68	8.70	1.71	6.61	12.99	21.01	23.09	21.21
PS_2113_a	<input checked="" type="checkbox"/>	0.07	5.03	5.95	1.34	7.87	15.75	22.14	21.82	20.03
PSA_2114	<input checked="" type="checkbox"/>	0.50	6.18	6.29	4.40	7.20	12.36	19.27	21.85	21.95

Percentages taken from the final laser data.

\* Final laser data not in percentages

- no laser analysis

**Table 3. Summary of the particle size information received from participating laboratories and replicate analysis laboratory for the fifty-fifth particle size distribution - PS55.**

**Benchmark Data**

Sample	% Gravel	% Sand	% Silt/Clay	Sediment Description (Post analysis)	Summary Data APEM verification			Sediment Description (Post analysis)
					% Gravel	% Sand	% Silt/Clay	
REPLICATE 1	80.2	3.7	16.1	Gravel	<input checked="" type="checkbox"/> 80.2	<input checked="" type="checkbox"/> 3.7	<input checked="" type="checkbox"/> 16.1	<input checked="" type="checkbox"/> Gravel
REPLICATE 2	79.6	3.7	16.7	Muddy Gravel	<input checked="" type="checkbox"/> 79.6	<input checked="" type="checkbox"/> 3.7	<input checked="" type="checkbox"/> 16.7	<input checked="" type="checkbox"/> Muddy Gravel
REPLICATE 3	80.0	3.7	16.3	Muddy Gravel	<input checked="" type="checkbox"/> 80.0	<input checked="" type="checkbox"/> 3.7	<input checked="" type="checkbox"/> 16.3	<input checked="" type="checkbox"/> Muddy Gravel
REPLICATE 4	80.0	3.9	16.1	Muddy Gravel	<input checked="" type="checkbox"/> 80.0	<input checked="" type="checkbox"/> 3.9	<input checked="" type="checkbox"/> 16.1	<input checked="" type="checkbox"/> Muddy Gravel
REPLICATE 5	80.2	3.7	16.1	Gravel	<input checked="" type="checkbox"/> 80.2	<input checked="" type="checkbox"/> 3.7	<input checked="" type="checkbox"/> 16.1	<input checked="" type="checkbox"/> Gravel
REP AVERAGE	80.0	3.7	16.3	Gravel	<input checked="" type="checkbox"/> 80.0	<input checked="" type="checkbox"/> 3.7	<input checked="" type="checkbox"/> 16.3	<input checked="" type="checkbox"/> Gravel

**Participant Data**


Lab	% Gravel	% Sand	% Silt/Clay	Sediment Description (Post analysis)	Summary Data APEM verification			Sediment Description (Post analysis)
					% Gravel	% Sand	% Silt/Clay	
PSA_2101	79.6	5.2	15.2	Sandy, muddy gravel	<input checked="" type="checkbox"/> 79.6	<input checked="" type="checkbox"/> 5.2	<input checked="" type="checkbox"/> 15.2	<input checked="" type="checkbox"/> Muddy Gravel
PSA_2102	79.8	3.0	17.2	Muddy gravel	<input checked="" type="checkbox"/> 79.8	<input checked="" type="checkbox"/> 3.0	<input checked="" type="checkbox"/> 17.2	<input checked="" type="checkbox"/> Muddy Gravel
PSA_2103	79.9	9.8	10.3	Muddy gravel	<input checked="" type="checkbox"/> 79.9	<input checked="" type="checkbox"/> 9.8	<input checked="" type="checkbox"/> 10.3	<input checked="" type="checkbox"/> Muddy Gravel
PSA_2105	80.2	3.7	16.1	Gravel	<input checked="" type="checkbox"/> 80.2	<input checked="" type="checkbox"/> 3.7	<input checked="" type="checkbox"/> 16.1	<input checked="" type="checkbox"/> Gravel
PSA_2106	81.3	5.4	13.3	Muddy gravel	<input checked="" type="checkbox"/> 81.3	<input checked="" type="checkbox"/> 5.8	<input checked="" type="checkbox"/> 12.9	<input checked="" type="checkbox"/> Gravel
PSA_2107	81.5	2.0	16.5	Gravel	<input checked="" type="checkbox"/> 81.5	<input checked="" type="checkbox"/> 2.0	<input checked="" type="checkbox"/> 16.5	<input checked="" type="checkbox"/> Gravel
PSA_2108	80.1	5.1	14.8	Gravel	<input checked="" type="checkbox"/> 80.1	<input checked="" type="checkbox"/> 5.1	<input checked="" type="checkbox"/> 14.8	<input checked="" type="checkbox"/> Gravel
PSA_2109	78.5	5.5	16.0	Muddy gravel	<input checked="" type="checkbox"/> 78.5	<input checked="" type="checkbox"/> 5.5	<input checked="" type="checkbox"/> 16.0	<input checked="" type="checkbox"/> Muddy Gravel
PSA_2110	-	-	-	-	<input checked="" type="checkbox"/> 80.5	<input checked="" type="checkbox"/> 2.8	<input checked="" type="checkbox"/> 16.7	<input checked="" type="checkbox"/> Gravel
PSA_2111	61.4	5.9	32.7	Muddy gravel	<input checked="" type="checkbox"/> 61.4	<input checked="" type="checkbox"/> 5.9	<input checked="" type="checkbox"/> 32.7	<input checked="" type="checkbox"/> Muddy Gravel
PSA_2112	80.3	2.1	17.6	Gravel	<input checked="" type="checkbox"/> 80.3	<input checked="" type="checkbox"/> 2.1	<input checked="" type="checkbox"/> 17.6	<input checked="" type="checkbox"/> Gravel
PSA_2113	80.1	3.0	16.9	Gravel (with fine silt)	<input checked="" type="checkbox"/> 80.1	<input checked="" type="checkbox"/> 3.0	<input checked="" type="checkbox"/> 16.9	<input checked="" type="checkbox"/> Gravel
PSA_2113_a	80.5	2.5	17.1	Gravel (with fine silt)	<input checked="" type="checkbox"/> 80.5	<input checked="" type="checkbox"/> 2.5	<input checked="" type="checkbox"/> 17.1	<input checked="" type="checkbox"/> Gravel
PSA_2114	79.90	3.90	16.20	Muddy gravel	<input checked="" type="checkbox"/> 79.9	<input checked="" type="checkbox"/> 3.5	<input checked="" type="checkbox"/> 16.6	<input checked="" type="checkbox"/> Muddy Gravel

- -Data not provided

- Participant calculations of % Gravel, % Sand and % Silt/Clay are correct based on the final data submitted

- Participant calculation different from APEM calculation using GRADISTAT on final merged data


**Figure 1. Particle size distribution curves resulting from analysis of 5 replicate samples of sediment distributed as PS55 (Benchmark Data).**








**Figure 3. Particle size ternary diagram for PS55, including the Benchmark replicates an all participating laboratories for Gravel, Sand and Mud.**



**Figure 4. Bar chart for PS55 showing the percentage gravel, sand, silt and mud recorded by each participating laboratory and the benchmark average.**



**Table 4. Summary of z-scores for each phi-interval for PS55; data from all participating laboratories included in mean and standard deviation calculations.**

	Phi-interval																			
	-6.50 to -6.00	-6.00 to -5.50	-5.50 to -5.00	-5.00 to -4.50	-4.50 to -4.00	-4.00 to -3.50	-3.50 to -3.00	-3.00 to -2.50	-2.50 to -2.00	-2.00 to -1.50	-1.50 to -1.00	-1.00 to -0.50	-0.50 to 0.00	0.00 to 0.50	0.50 to 1.00	1.00 to 1.50	1.50 to 2.00	2.00 to 2.50	2.50 to 3.00	3.00 to 3.50
Benchmark Average	0.00	0.00	0.00	0.00	0.00	-0.23	0.26	-0.05	0.45	0.02	-0.45	-0.52	-0.53	-0.53	-0.63	-0.90	-0.09	0.73	0.13	0.43
PSA_2101	0.00	0.00	0.00	0.00	0.00	-0.73	-0.74	0.75	1.88	1.00	-0.54	-0.38	-0.03	-0.28	-0.24	-0.04	0.02	0.82	1.59	1.60
PSA_2102	0.00	0.00	0.00	0.00	0.00	-0.91	0.13	-0.37	2.15	1.08	0.91	2.20	1.30	0.21	-0.16	-0.21	-0.43	-0.58	-0.96	-1.38
PSA_2103	0.00	0.00	0.00	0.00	0.00	-0.56	0.16	0.56	-0.48	-0.28	-0.55	-0.52	-0.29	-0.53	1.22	2.52	2.57	2.21	1.97	1.45
PSA_2105	0.00	0.00	0.00	0.00	0.00	-0.54	0.62	-0.40	0.70	0.16	-0.54	-0.52	-0.53	-0.53	-0.63	-0.89	-0.02	0.49	0.07	0.40
PSA_2106	0.00	0.00	0.00	0.00	0.00	0.17	0.15	0.61	-0.18	1.59	0.06	-0.52	-0.46	3.07	2.98	1.75	-0.45	-0.74	-1.28	0.62
PSA_2107	0.00	0.00	0.00	0.00	0.00	-0.52	-0.30	1.48	0.00	0.33	0.06	-0.52	-0.53	-0.53	-0.62	-0.71	-1.13	-1.36	-1.06	-0.13
PSA_2108	0.00	0.00	0.00	0.00	0.00	-0.91	0.46	0.19	0.07	-1.91	-0.55	-0.41	-0.29	0.05	0.41	0.52	0.28	0.09	0.31	0.33
PSA_2109	0.00	0.00	0.00	0.00	0.00	-0.07	-0.13	0.22	-0.31	-1.52	0.28	-0.52	-0.53	0.35	-0.55	0.00	1.80	1.18	0.84	-1.28
PSA_2110	0.00	0.00	0.00	0.00	0.00	0.08	-0.05	0.92	-0.86	-0.19	-0.61	0.85	-0.53	-0.53	-0.34	-0.36	-0.74	-1.05	-0.79	-0.02
PSA_2111	0.00	0.00	0.00	0.00	0.00	-0.91	-2.73	-1.83	-0.54	0.18	-0.61	-0.52	-0.53	1.35	0.74	0.67	0.46	0.17	0.82	0.81
PSA_2112	0.00	0.00	0.00	0.00	0.00	1.15	-0.52	0.59	0.20	0.90	-0.27	-0.38	-0.28	-0.53	-0.59	-0.68	-1.06	-1.36	-0.98	-0.34
PSA_2113	0.00	0.00	0.00	0.00	0.00	2.83	0.97	-1.53	-1.23	-1.49	-0.48	-0.11	0.67	-0.53	-0.63	-0.82	-0.33	0.33	0.32	-1.18
PSA_2113_a	0.00	0.00	0.00	0.00	0.00	0.52	1.97	-1.72	-1.43	-0.04	3.26	2.41	3.08	-0.53	-0.59	-0.62	-0.46	-0.41	-0.83	-1.64
PSA_2114	0.00	0.00	0.00	0.00	0.00	0.62	-0.24	0.56	-0.41	0.16	0.04	-0.52	-0.53	-0.53	-0.36	-0.25	-0.42	-0.52	-0.14	0.35
$\bar{x}$	0.00	0.00	0.00	0.00	0.00	1.11	42.45	30.99	4.01	0.341	0.009	0.004	0.002	0.020	0.236	0.615	1.037	0.968	0.551	0.35
s	0.00	0.00	0.00	0.00	0.00	1.22	3.83	2.92	1.16	0.058	0.015	0.007	0.004	0.037	0.375	0.629	0.617	0.437	0.217	0.16

z < -1.96 or z > 1.96  
 All laboratories recorded zero therefore mean and standard deviation equal zero.

**Table 4. Summary of z-scores for each phi-interval for PS55; data from all participating laboratories included in mean and standard deviation calculations.**


	Phi-interval																			
	3.50 to 4.00	4.00 to 4.50	4.50 to 5.00	5.00 to 5.50	5.50 to 6.00	6.00 to 6.50	6.50 to 7.00	7.00 to 7.50	7.50 to 8.00	8.00 to 8.50	8.50 to 9.00	9.00 to 9.50	9.50 to 10.00	10.00 to 10.50	10.50 to 11.00	11.00 to 11.50	11.50 to 12.00	12.00 to 12.50	12.50 to 13.00	13.00 to 13.50
Benchmark Average	-0.20	-0.40	-0.51	-0.23	-0.55	-0.45	-0.26	-0.17	-0.42	-0.22	-0.05	0.08	-0.25	-0.31	1.06	2.12	2.27	2.29	2.29	2.13
PSA_2101	0.89	-0.33	-0.28	-0.42	-0.57	-0.28	-0.31	-0.31	-0.48	-0.36	-0.21	0.15	0.42	0.21	0.80	1.00	0.95	0.92	0.95	1.49
PSA_2102	-0.75	-0.30	0.34	0.34	-0.03	0.41	0.36	0.37	-0.09	0.60	0.64	0.18	-0.84	-1.26	-1.00	-0.69	-0.46	-0.46	-0.46	-0.48
PSA_2103	2.12	0.38	1.50	0.44	-0.63	-0.88	-1.20	-1.27	-0.97	-1.34	-1.26	-0.99	-0.62	-0.52	-0.33	-0.50	-0.46	-0.46	-0.46	-0.48
PSA_2105	-0.17	-0.41	-0.54	-0.25	-0.57	-0.49	-0.33	-0.23	-0.44	-0.24	-0.05	0.11	-0.23	-0.29	1.10	2.17	2.32	2.32	2.30	2.12
PSA_2106	0.14	-0.11	0.64	0.68	-0.53	-0.22	-0.63	-1.04	-0.66	-0.58	-0.26	-0.27	0.52	-0.24	-0.69	-0.69	-0.46	-0.46	-0.46	-0.48
PSA_2107	0.03	-0.32	-0.25	-0.29	-0.42	0.13	0.20	0.30	-0.18	0.19	0.23	0.35	0.29	0.22	0.35	-0.16	-0.46	-0.46	-0.46	-0.48
PSA_2108	0.16	-0.27	0.08	-0.07	-0.42	-0.13	-0.18	-0.17	-0.39	-0.13	-0.10	0.01	0.21	0.06	0.20	-0.31	-0.46	-0.46	-0.46	-0.48
PSA_2109	-1.04	3.49	-2.61	-2.67	3.10	-2.34	-2.24	-2.04	3.30	-2.11	-2.14	-1.94	-1.55	-1.26	-1.00	-0.69	-0.46	-0.46	-0.46	-0.48
PSA_2110	0.11	-0.33	-0.15	-0.05	-0.18	0.30	0.42	0.37	-0.17	0.22	0.03	-0.21	0.14	0.95	-1.00	-0.69	-0.46	-0.46	-0.46	-0.48
PSA_2111	1.54	0.31	1.89	2.33	1.46	2.49	2.39	2.36	0.86	2.36	2.64	2.97	2.76	2.46	2.28	0.56	-0.46	-0.46	-0.46	-0.48
PSA_2112	0.03	-0.40	-0.03	-0.09	-0.25	0.32	0.54	0.79	0.20	0.99	0.50	-0.32	-1.41	-1.26	-1.00	-0.69	-0.46	-0.46	-0.46	-0.48
PSA_2113	-1.49	-0.53	-0.15	-0.06	-0.22	0.40	0.47	0.44	-0.14	0.21	0.03	-0.05	0.13	0.06	0.22	-0.33	-0.46	-0.46	-0.46	-0.48
PSA_2113_a	-1.53	-0.44	0.28	0.48	0.13	0.62	0.44	0.27	-0.24	0.04	-0.07	-0.13	-0.02	-0.12	0.03	-0.43	-0.46	-0.46	-0.46	-0.48
PSA_2114	0.17	-0.35	-0.21	-0.12	-0.32	0.12	0.33	0.32	-0.19	0.34	0.07	0.06	0.45	1.30	-1.00	-0.69	-0.46	-0.46	-0.46	-0.48
$\bar{x}$	0.440	0.907	0.891	1.154	1.616	1.620	1.893	1.997	2.427	1.464	0.979	0.572	0.349	0.325	0.217	0.164	0.118	0.086	0.055	0.030
s	0.179	0.815	0.342	0.432	0.703	0.692	0.846	0.977	1.831	0.693	0.456	0.295	0.225	0.258	0.216	0.240	0.255	0.187	0.118	0.063





z < -1.96 or z > 1.96

All laboratories recorded zero therefore mean and standard deviation equal zero.



Figure 5. Summary of z-scores at each half phi-interval for the Benchmark Average for PS55.



**Figure 6. Summary of z-scores in each half phi-interval for PS55 for a) PSA\_2101 and PSA\_2102 and b) PSA\_2103 and PSA\_2105**



**Figure 7. Summary of z-scores in each half phi-interval for PS55 for a) PSA\_2106 and PSA\_2107 and b) PSA\_2108 and PSA\_2109**





**Figure 8. Summary of z-scores in each half phi-interval for PS55 for a) PSA\_2110, PSA\_2111 and PSA\_2112 and b) PSA\_2113, PSA\_2113\_a and PSA\_2114**






Figure 9. Cluster dendrogram of PS55 including all participating laboratories, with the benchmark replicates averaged.



## Results of SIMPROF testing on PSA Ring test PS55 data

Data was entered into PRIMER v.6.1.13 in half-phi intervals; any missing data was entered as zero. The data did not need to be transformed as all data was on a similar percentage scale. A Euclidean distance matrix was created from the data. The Euclidean distance between two samples  $j$  and  $k$ , is defined algebraically as;

$$d_{jk} = \sqrt{\sum_{i=1}^p (y_{ij} - y_{ik})^2}$$

From this distance matrix cluster analysis was carried out. The most commonly used clustering methods are hierarchical agglomerative methods. These use a similarity or distance matrix as their starting point and successively fuse the samples into groups and the groups into larger clusters, starting with the highest mutual similarities then gradually lowering the similarity level at which groups are formed. The result of a hierarchical clustering is represented by a dendrogram (Figure 10), with the X axis representing the full set of samples (labs) and the y axis defining the similarity at which two samples or groups are considered to have fused (Clarke & Warwick, 2001)<sup>1</sup>.

A similarity profile routine, SIMPROF was run on the data in conjunction with the cluster analysis. Similarity profile analysis examines whether the similarities observed in the data are smaller and/or larger than those expected by chance. The red SIMPROF lines on the dendrogram (Figure 10) indicate that labs cannot be distinguished from each other at the 5% significance level; black lines indicate labs that can be distinguished from each other.

The cluster analysis separates the labs into 6 SIMPROF cluster groups, 2 of these groups comprise a single lab.

Cluster group A consists of one lab, PSA\_2111, Figure 4 and Table 3 show that this laboratory recorded a higher percentage of mud (32.7%) and a lower percentage of gravel (61.4%) compared to other laboratories. This can also be seen on the cumulative percentage graph in Figure 2. This could be due to the large amount of sediment less than 1mm recorded (503.8g), approximately 200g more than other labs.

Cluster group C consists of one laboratory, PSA\_2109. This laboratory used alternate methods and did not provide data in half phi intervals; this can be seen in the cumulative percentage graph in Figure 2. Despite this the percentage gravel, sand and mud recorded was similar to the majority of other laboratories.

The particle size ternary diagram in Figure 3 shows that the laboratories fell into 3 sediment classifications; Slightly sandy muddy Gravel (PSA\_2111), very slightly sandy slightly muddy Gravel (Benchmark, PSA\_2102, PSA\_2105, PSA\_2107, PSA\_2110, PSA\_2112, PSA\_2113, PSA\_2113\_a and PSA\_2114) and Slightly sandy slightly muddy Gravel (PSA\_2101, PSA\_2103, PSA\_22106, PSA\_2108 and PSA\_2109).

---

<sup>1</sup> Clarke, K.R. & Warwick, R.M. (2001). Changes in Marine Communities; an approach to statistical analysis and interpretation. 2<sup>nd</sup> Ed. PRIMER-E Ltd. Plymouth.

## Appendices

**NMBAQCS - PS Exercise Data Workbook (Page 2 -  
Final Merged Data Submission)**

Return to APEM Ltd. by 06-02-15

Exercise Code:	<b>PS55</b>
LabCode:	<b>PSA_2101</b>
Sample Code:	<b>PS552101</b>

Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.00
-6.00 to -5.50; 45 mm	0.00
-5.50 to -5.00; 31.5 mm	0.00
-5.00 to -4.50; 22.4 mm	0.00
-4.50 to -4.00; 16 mm	0.00
-4.00 to -3.50; 11.2 mm	0.22
-3.50 to -3.00; 8 mm	39.62
-3.00 to -2.50; 5.6 mm	33.18
-2.50 to -2.00; 4 mm	6.19
-2.00 to -1.50; 2.8 mm	0.40
-1.50 to -1.00; 2 mm	0.00
-1.00 to -0.50; 1.4 mm	0.00
-0.50 to 0.00; 1 mm	0.00
0.00 to 0.50; (707 µm)	0.01
0.50 to 1.00; (500 µm)	0.15
1.00 to 1.50; (353.6 µm)	0.59
1.50 to 2.00; (250 µm)	1.05
2.00 to 2.50; (176.8 µm)	1.33
2.50 to 3.00; (125 µm)	0.90
3.00 to 3.50; (88.39 µm)	0.61
3.50 to 4.00; (62.5 µm)	0.60
4.00 to 4.50; (44.19 µm)	0.64
4.50 to 5.00; (31.25 µm)	0.80
5.00 to 5.50; (22.097 µm)	0.97
5.50 to 6.00; (15.625 µm)	1.21
6.00 to 6.50; (11.049 µm)	1.43
6.50 to 7.00; (7.813 µm)	1.63
7.00 to 7.50; (5.524 µm)	1.70
7.50 to 8.00; (3.906 µm)	1.54
8.00 to 8.50; (2.762 µm)	1.22
8.50 to 9.00; (1.953 µm)	0.88
9.00 to 9.50; (1.381 µm)	0.62
9.50 to 10.00; (0.977 µm)	0.44
10.00 to 10.50; (0.691 µm)	0.38
10.50 to 11.00; (0.488 µm)	0.39
11.00 to 11.50; (0.345 µm)	0.40
11.50 to 12.00; (0.244 µm)	0.36
12.00 to 12.50; (0.173 µm)	0.26
12.50 to 13.00; (0.122 µm)	0.17
13.00 to 13.50; (0.086 µm)	0.12

**NMBAQCS - PS Exercise Data Workbook (Page 2 -  
Final Merged Data Submission)**

Return to APEM Ltd. by 06-02-15

Exercise Code:	<b>PS55</b>
LabCode:	<b>PSA_2102</b>
Sample Code:	<b>PS552102</b>

Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	0.0000
-4.00 to -3.50; 11.2 mm	0.0000
-3.50 to -3.00; 8 mm	42.9461
-3.00 to -2.50; 5.6 mm	29.9209
-2.50 to -2.00; 4 mm	6.4972
-2.00 to -1.50; 2.8 mm	0.4035
-1.50 to -1.00; 2 mm	0.0231
-1.00 to -0.50; 1.4 mm	0.0198
-0.50 to 0.00; 1 mm	0.0077
0.00 to 0.50; (707 µm)	0.0277
0.50 to 1.00; (500 µm)	0.1759
1.00 to 1.50; (353.6 µm)	0.4855
1.50 to 2.00; (250 µm)	0.7703
2.00 to 2.50; (176.8 µm)	0.7139
2.50 to 3.00; (125 µm)	0.3429
3.00 to 3.50; (88.39 µm)	0.1351
3.50 to 4.00; (62.5 µm)	0.3066
4.00 to 4.50; (44.19 µm)	0.6661
4.50 to 5.00; (31.25 µm)	1.0072
5.00 to 5.50; (22.097 µm)	1.2999
5.50 to 6.00; (15.625 µm)	1.5914
6.00 to 6.50; (11.049 µm)	1.9031
6.50 to 7.00; (7.813 µm)	2.1940
7.00 to 7.50; (5.524 µm)	2.3565
7.50 to 8.00; (3.906 µm)	2.2681
8.00 to 8.50; (2.762 µm)	1.8805
8.50 to 9.00; (1.953 µm)	1.2708
9.00 to 9.50; (1.381 µm)	0.6264
9.50 to 10.00; (0.977 µm)	0.1599
10.00 to 10.50; (0.691 µm)	0.0000
10.50 to 11.00; (0.488 µm)	0.0000
11.00 to 11.50; (0.345 µm)	0.0000
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000

**NMBAQCS - PS Exercise Data Workbook (Page 2 -  
Final Merged Data Submission)**

Return to APEM Ltd. by 06-02-15

Exercise Code:	<b>PS55</b>
LabCode:	<b>PSA_2103</b>
Sample Code:	<b>PS552103</b>

Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	0.0000
-4.00 to -3.50; 11.2 mm	4.3100
-3.50 to -3.00; 8 mm	428.5400
-3.00 to -2.50; 5.6 mm	324.8000
-2.50 to -2.00; 4 mm	34.3100
-2.00 to -1.50; 2.8 mm	3.2300
-1.50 to -1.00; 2 mm	0.0100
-1.00 to -0.50; 1.4 mm	0.0000
-0.50 to 0.00; 1 mm	0.0100
0.00 to 0.50; (707 µm)	0.0000
0.50 to 1.00; (500 µm)	6.9172
1.00 to 1.50; (353.6 µm)	21.8979
1.50 to 2.00; (250 µm)	26.1362
2.00 to 2.50; (176.8 µm)	19.2590
2.50 to 3.00; (125 µm)	9.7494
3.00 to 3.50; (88.39 µm)	5.8043
3.50 to 4.00; (62.5 µm)	8.1434
4.00 to 4.50; (44.19 µm)	12.1351
4.50 to 5.00; (31.25 µm)	13.9677
5.00 to 5.50; (22.097 µm)	13.3547
5.50 to 6.00; (15.625 µm)	11.6953
6.00 to 6.50; (11.049 µm)	10.0760
6.50 to 7.00; (7.813 µm)	8.7298
7.00 to 7.50; (5.524 µm)	7.5703
7.50 to 8.00; (3.906 µm)	6.4441
8.00 to 8.50; (2.762 µm)	5.3379
8.50 to 9.00; (1.953 µm)	4.0317
9.00 to 9.50; (1.381 µm)	2.7989
9.50 to 10.00; (0.977 µm)	2.0858
10.00 to 10.50; (0.691 µm)	1.9059
10.50 to 11.00; (0.488 µm)	1.4594
11.00 to 11.50; (0.345 µm)	0.4465
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000

**NMBAQCS - PS Exercise Data Workbook (Page 2 -  
Final Merged Data Submission)**

Return to APEM Ltd. by 06-02-15

Exercise Code:	<b>PS55</b>
LabCode:	<b>PSA_2105</b>
Sample Code:	<b>PS552105</b>

Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	0.0000
-4.00 to -3.50; 11.2 mm	0.4590
-3.50 to -3.00; 8 mm	44.8010
-3.00 to -2.50; 5.6 mm	29.8093
-2.50 to -2.00; 4 mm	4.8226
-2.00 to -1.50; 2.8 mm	0.3501
-1.50 to -1.00; 2 mm	0.0011
-1.00 to -0.50; 1.4 mm	0.0000
-0.50 to 0.00; 1 mm	0.0000
0.00 to 0.50; (707 µm)	0.0000
0.50 to 1.00; (500 µm)	0.0000
1.00 to 1.50; (353.6 µm)	0.0577
1.50 to 2.00; (250 µm)	1.0225
2.00 to 2.50; (176.8 µm)	1.1812
2.50 to 3.00; (125 µm)	0.5664
3.00 to 3.50; (88.39 µm)	0.4177
3.50 to 4.00; (62.5 µm)	0.4099
4.00 to 4.50; (44.19 µm)	0.5743
4.50 to 5.00; (31.25 µm)	0.7050
5.00 to 5.50; (22.097 µm)	1.0447
5.50 to 6.00; (15.625 µm)	1.2158
6.00 to 6.50; (11.049 µm)	1.2802
6.50 to 7.00; (7.813 µm)	1.6170
7.00 to 7.50; (5.524 µm)	1.7727
7.50 to 8.00; (3.906 µm)	1.6244
8.00 to 8.50; (2.762 µm)	1.3001
8.50 to 9.00; (1.953 µm)	0.9580
9.00 to 9.50; (1.381 µm)	0.6042
9.50 to 10.00; (0.977 µm)	0.2975
10.00 to 10.50; (0.691 µm)	0.2492
10.50 to 11.00; (0.488 µm)	0.4546
11.00 to 11.50; (0.345 µm)	0.6842
11.50 to 12.00; (0.244 µm)	0.7090
12.00 to 12.50; (0.173 µm)	0.5207
12.50 to 13.00; (0.122 µm)	0.3259
13.00 to 13.50; (0.086 µm)	0.1640



**NMBAQCS - PS Exercise Data Workbook (Page 2 -  
Final Merged Data Submission)**

Return to APEM Ltd. by 06-02-15

Exercise Code:	<b>PS55</b>
LabCode:	<b>PSA_2106</b>
Sample Code:	<b>PS552106</b>

Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	0.0000
-4.00 to -3.50; 11.2 mm	1.3142
-3.50 to -3.00; 8 mm	43.0080
-3.00 to -2.50; 5.6 mm	32.7777
-2.50 to -2.00; 4 mm	3.7978
-2.00 to -1.50; 2.8 mm	0.4333
-1.50 to -1.00; 2 mm	0.0102
-1.00 to -0.50; 1.4 mm	0.0000
-0.50 to 0.00; 1 mm	0.0003
0.00 to 0.50; (707 µm)	0.1343
0.50 to 1.00; (500 µm)	1.3528
1.00 to 1.50; (353.6 µm)	1.7186
1.50 to 2.00; (250 µm)	0.7595
2.00 to 2.50; (176.8 µm)	0.6438
2.50 to 3.00; (125 µm)	0.2743
3.00 to 3.50; (88.39 µm)	0.4516
3.50 to 4.00; (62.5 µm)	0.4646
4.00 to 4.50; (44.19 µm)	0.8173
4.50 to 5.00; (31.25 µm)	1.1084
5.00 to 5.50; (22.097 µm)	1.4480
5.50 to 6.00; (15.625 µm)	1.2428
6.00 to 6.50; (11.049 µm)	1.4667
6.50 to 7.00; (7.813 µm)	1.3622
7.00 to 7.50; (5.524 µm)	0.9796
7.50 to 8.00; (3.906 µm)	1.2148
8.00 to 8.50; (2.762 µm)	1.0636
8.50 to 9.00; (1.953 µm)	0.8584
9.00 to 9.50; (1.381 µm)	0.4926
9.50 to 10.00; (0.977 µm)	0.4665
10.00 to 10.50; (0.691 µm)	0.2631
10.50 to 11.00; (0.488 µm)	0.0685
11.00 to 11.50; (0.345 µm)	0.0000
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000

**NMBAQCS - PS Exercise Data Workbook (Page 2 -  
Final Merged Data Submission)**

Return to APEM Ltd. by 06-02-15

Exercise Code:	<b>PS55</b>
LabCode:	<b>PSA_2107</b>
Sample Code:	<b>PS552107</b>

Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	0.0000
-4.00 to -3.50; 11.2 mm	0.4837
-3.50 to -3.00; 8 mm	41.2824
-3.00 to -2.50; 5.6 mm	35.3232
-2.50 to -2.00; 4 mm	4.0140
-2.00 to -1.50; 2.8 mm	0.3602
-1.50 to -1.00; 2 mm	0.0103
-1.00 to -0.50; 1.4 mm	0.0000
-0.50 to 0.00; 1 mm	0.0000
0.00 to 0.50; (707 µm)	0.0000
0.50 to 1.00; (500 µm)	0.0023
1.00 to 1.50; (353.6 µm)	0.1679
1.50 to 2.00; (250 µm)	0.3399
2.00 to 2.50; (176.8 µm)	0.3759
2.50 to 3.00; (125 µm)	0.3202
3.00 to 3.50; (88.39 µm)	0.3328
3.50 to 4.00; (62.5 µm)	0.4459
4.00 to 4.50; (44.19 µm)	0.6419
4.50 to 5.00; (31.25 µm)	0.8061
5.00 to 5.50; (22.097 µm)	1.0282
5.50 to 6.00; (15.625 µm)	1.3199
6.00 to 6.50; (11.049 µm)	1.7091
6.50 to 7.00; (7.813 µm)	2.0579
7.00 to 7.50; (5.524 µm)	2.2869
7.50 to 8.00; (3.906 µm)	2.1048
8.00 to 8.50; (2.762 µm)	1.5988
8.50 to 9.00; (1.953 µm)	1.0852
9.00 to 9.50; (1.381 µm)	0.6744
9.50 to 10.00; (0.977 µm)	0.4153
10.00 to 10.50; (0.691 µm)	0.3823
10.50 to 11.00; (0.488 µm)	0.2937
11.00 to 11.50; (0.345 µm)	0.1266
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000

**NMBAQCS - PS Exercise Data Workbook (Page 2 -  
Final Merged Data Submission)**

Return to APEM Ltd. by 06-02-15

Exercise Code:	<b>PS55</b>
LabCode:	<b>PSA_2108</b>
Sample Code:	<b>PS552108</b>

Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	0.0000
-4.00 to -3.50; 11.2 mm	0.0000
-3.50 to -3.00; 8 mm	44.2133
-3.00 to -2.50; 5.6 mm	31.5525
-2.50 to -2.00; 4 mm	4.0961
-2.00 to -1.50; 2.8 mm	0.2305
-1.50 to -1.00; 2 mm	0.0010
-1.00 to -0.50; 1.4 mm	0.0008
-0.50 to 0.00; 1 mm	0.0010
0.00 to 0.50; (707 µm)	0.0215
0.50 to 1.00; (500 µm)	0.3882
1.00 to 1.50; (353.6 µm)	0.9451
1.50 to 2.00; (250 µm)	1.2082
2.00 to 2.50; (176.8 µm)	1.0068
2.50 to 3.00; (125 µm)	0.6175
3.00 to 3.50; (88.39 µm)	0.4055
3.50 to 4.00; (62.5 µm)	0.4699
4.00 to 4.50; (44.19 µm)	0.6866
4.50 to 5.00; (31.25 µm)	0.9197
5.00 to 5.50; (22.097 µm)	1.1238
5.50 to 6.00; (15.625 µm)	1.3189
6.00 to 6.50; (11.049 µm)	1.5312
6.50 to 7.00; (7.813 µm)	1.7368
7.00 to 7.50; (5.524 µm)	1.8353
7.50 to 8.00; (3.906 µm)	1.7179
8.00 to 8.50; (2.762 µm)	1.3743
8.50 to 9.00; (1.953 µm)	0.9350
9.00 to 9.50; (1.381 µm)	0.5766
9.50 to 10.00; (0.977 µm)	0.3969
10.00 to 10.50; (0.691 µm)	0.3393
10.50 to 11.00; (0.488 µm)	0.2604
11.00 to 11.50; (0.345 µm)	0.0892
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000

**NMBAQCS - PS Exercise Data Workbook (Page 2 -  
Final Merged Data Submission)**

Return to APEM Ltd. by 06-02-15

Exercise Code:	<b>PS55</b>
LabCode:	<b>PSA_2109</b>
Sample Code:	<b>PS552109</b>

Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	0.0000
-4.00 to -3.50; 11.2 mm	10.4190
-3.50 to -3.00; 8 mm	424.0460
-3.00 to -2.50; 5.6 mm	320.0310
-2.50 to -2.00; 4 mm	36.8830
-2.00 to -1.50; 2.8 mm	2.5610
-1.50 to -1.00; 2 mm	0.1370
-1.00 to -0.50; 1.4 mm	0.0000
-0.50 to 0.00; 1 mm	0.0000
0.00 to 0.50; (707 µm)	0.3320
0.50 to 1.00; (500 µm)	0.3200
1.00 to 1.50; (353.6 µm)	6.1880
1.50 to 2.00; (250 µm)	21.7360
2.00 to 2.50; (176.8 µm)	15.0100
2.50 to 3.00; (125 µm)	7.4120
3.00 to 3.50; (88.39 µm)	1.5210
3.50 to 4.00; (62.5 µm)	2.5760
4.00 to 4.50; (44.19 µm)	37.9160
4.50 to 5.00; (31.25 µm)	0.0000
5.00 to 5.50; (22.097 µm)	0.0000
5.50 to 6.00; (15.625 µm)	38.4021
6.00 to 6.50; (11.049 µm)	0.0000
6.50 to 7.00; (7.813 µm)	0.0000
7.00 to 7.50; (5.524 µm)	0.0000
7.50 to 8.00; (3.906 µm)	85.7160
8.00 to 8.50; (2.762 µm)	0.0000
8.50 to 9.00; (1.953 µm)	0.0000
9.00 to 9.50; (1.381 µm)	0.0000
9.50 to 10.00; (0.977 µm)	0.0000
10.00 to 10.50; (0.691 µm)	0.0000
10.50 to 11.00; (0.488 µm)	0.0000
11.00 to 11.50; (0.345 µm)	0.0000
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000

**NMBAQCS - PS Exercise Data Workbook (Page 2 -  
Final Merged Data Submission)**

Return to APEM Ltd. by 06-02-15

Exercise Code:	<b>PS55</b>
LabCode:	<b>PSA_2110</b>
Sample Code:	<b>PS552110</b>

Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	0.0000
-4.00 to -3.50; 11.2 mm	1.2100
-3.50 to -3.00; 8 mm	42.2600
-3.00 to -2.50; 5.6 mm	33.6800
-2.50 to -2.00; 4 mm	3.0100
-2.00 to -1.50; 2.8 mm	0.3300
-1.50 to -1.00; 2 mm	0.0000
-1.00 to -0.50; 1.4 mm	0.0100
-0.50 to 0.00; 1 mm	0.0000
0.00 to 0.50; (707 µm)	0.0000
0.50 to 1.00; (500 µm)	0.1100
1.00 to 1.50; (353.6 µm)	0.3900
1.50 to 2.00; (250 µm)	0.5800
2.00 to 2.50; (176.8 µm)	0.5100
2.50 to 3.00; (125 µm)	0.3800
3.00 to 3.50; (88.39 µm)	0.3500
3.50 to 4.00; (62.5 µm)	0.4600
4.00 to 4.50; (44.19 µm)	0.6400
4.50 to 5.00; (31.25 µm)	0.8400
5.00 to 5.50; (22.097 µm)	1.1300
5.50 to 6.00; (15.625 µm)	1.4900
6.00 to 6.50; (11.049 µm)	1.8300
6.50 to 7.00; (7.813 µm)	2.2500
7.00 to 7.50; (5.524 µm)	2.3600
7.50 to 8.00; (3.906 µm)	2.1200
8.00 to 8.50; (2.762 µm)	1.6200
8.50 to 9.00; (1.953 µm)	0.9900
9.00 to 9.50; (1.381 µm)	0.5100
9.50 to 10.00; (0.977 µm)	0.3800
10.00 to 10.50; (0.691 µm)	
10.50 to 11.00; (0.488 µm)	
11.00 to 11.50; (0.345 µm)	
11.50 to 12.00; (0.244 µm)	0.5700
12.00 to 12.50; (0.173 µm)	
12.50 to 13.00; (0.122 µm)	
13.00 to 13.50; (0.086 µm)	

**NMBAQCS - PS Exercise Data Workbook (Page 2 -  
Final Merged Data Submission)**

Return to APEM Ltd. by 06-02-15

Exercise Code:	<b>PS55</b>
LabCode:	<b>PSA_2111</b>
Sample Code:	<b>PS552111</b>

Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	0.0000
-4.00 to -3.50; 11.2 mm	0.0000
-3.50 to -3.00; 8 mm	417.1000
-3.00 to -2.50; 5.6 mm	334.3667
-2.50 to -2.00; 4 mm	44.0667
-2.00 to -1.50; 2.8 mm	4.5833
-1.50 to -1.00; 2 mm	0.0000
-1.00 to -0.50; 1.4 mm	0.0000
-0.50 to 0.00; 1 mm	0.0000
0.00 to 0.50; (707 µm)	0.9159
0.50 to 1.00; (500 µm)	6.7066
1.00 to 1.50; (353.6 µm)	13.5306
1.50 to 2.00; (250 µm)	17.2058
2.00 to 2.50; (176.8 µm)	13.5945
2.50 to 3.00; (125 µm)	9.5102
3.00 to 3.50; (88.39 µm)	6.2874
3.50 to 4.00; (62.5 µm)	9.3359
4.00 to 4.50; (44.19 µm)	15.1528
4.50 to 5.00; (31.25 µm)	20.0140
5.00 to 5.50; (22.097 µm)	28.1866
5.50 to 6.00; (15.625 µm)	34.4589
6.00 to 6.50; (11.049 µm)	43.6170
6.50 to 7.00; (7.813 µm)	51.0254
7.00 to 7.50; (5.524 µm)	56.1354
7.50 to 8.00; (3.906 µm)	52.0687
8.00 to 8.50; (2.762 µm)	40.4370
8.50 to 9.00; (1.953 µm)	28.4597
9.00 to 9.50; (1.381 µm)	18.8588
9.50 to 10.00; (0.977 µm)	12.6519
10.00 to 10.50; (0.691 µm)	12.5003
10.50 to 11.00; (0.488 µm)	9.2669
11.00 to 11.50; (0.345 µm)	3.8777
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000

**NMBAQCS - PS Exercise Data Workbook (Page 2 -  
Final Merged Data Submission)**

Return to APEM Ltd. by 06-02-15

Exercise Code:	<b>PS55</b>
LabCode:	<b>PSA_2112</b>
Sample Code:	<b>PS552112</b>

Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	0.0000
-4.00 to -3.50; 11.2 mm	24.4000
-3.50 to -3.00; 8 mm	391.8200
-3.00 to -2.50; 5.6 mm	317.0000
-2.50 to -2.00; 4 mm	41.1200
-2.00 to -1.50; 2.8 mm	3.8100
-1.50 to -1.00; 2 mm	0.0500
-1.00 to -0.50; 1.4 mm	0.0100
-0.50 to 0.00; 1 mm	0.0100
0.00 to 0.50; (707 µm)	0.0000
0.50 to 1.00; (500 µm)	0.1649
1.00 to 1.50; (353.6 µm)	1.8280
1.50 to 2.00; (250 µm)	3.6923
2.00 to 2.50; (176.8 µm)	3.6127
2.50 to 3.00; (125 µm)	3.2761
3.00 to 3.50; (88.39 µm)	2.8992
3.50 to 4.00; (62.5 µm)	4.3188
4.00 to 4.50; (44.19 µm)	5.6081
4.50 to 5.00; (31.25 µm)	8.5231
5.00 to 5.50; (22.097 µm)	10.7807
5.50 to 6.00; (15.625 µm)	13.9298
6.00 to 6.50; (11.049 µm)	17.8600
6.50 to 7.00; (7.813 µm)	22.7650
7.00 to 7.50; (5.524 µm)	26.8226
7.50 to 8.00; (3.906 µm)	26.9813
8.00 to 8.50; (2.762 µm)	20.8503
8.50 to 9.00; (1.953 µm)	11.6842
9.00 to 9.50; (1.381 µm)	4.6314
9.50 to 10.00; (0.977 µm)	0.3095
10.00 to 10.50; (0.691 µm)	0.0000
10.50 to 11.00; (0.488 µm)	0.0010
11.00 to 11.50; (0.345 µm)	0.0008
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000

**NMBAQCS - PS Exercise Data Workbook (Page 2 -  
Final Merged Data Submission)**

Return to APEM Ltd. by 06-02-15

Exercise Code:	<b>PS55</b>
LabCode:	<b>PSA_2113</b>
Sample Code:	<b>PS552113</b>

Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	0.0000
-4.00 to -3.50; 11.2 mm	4.5688
-3.50 to -3.00; 8 mm	46.1408
-3.00 to -2.50; 5.6 mm	26.5081
-2.50 to -2.00; 4 mm	2.5856
-2.00 to -1.50; 2.8 mm	0.2547
-1.50 to -1.00; 2 mm	0.0020
-1.00 to -0.50; 1.4 mm	0.0030
-0.50 to 0.00; 1 mm	0.0051
0.00 to 0.50; (707 µm)	0.0000
0.50 to 1.00; (500 µm)	0.0001
1.00 to 1.50; (353.6 µm)	0.1013
1.50 to 2.00; (250 µm)	0.8308
2.00 to 2.50; (176.8 µm)	1.1139
2.50 to 3.00; (125 µm)	0.6199
3.00 to 3.50; (88.39 µm)	0.1671
3.50 to 4.00; (62.5 µm)	0.1740
4.00 to 4.50; (44.19 µm)	0.4781
4.50 to 5.00; (31.25 µm)	0.8390
5.00 to 5.50; (22.097 µm)	1.1285
5.50 to 6.00; (15.625 µm)	1.4602
6.00 to 6.50; (11.049 µm)	1.8951
6.50 to 7.00; (7.813 µm)	2.2930
7.00 to 7.50; (5.524 µm)	2.4293
7.50 to 8.00; (3.906 µm)	2.1738
8.00 to 8.50; (2.762 µm)	1.6101
8.50 to 9.00; (1.953 µm)	0.9902
9.00 to 9.50; (1.381 µm)	0.5562
9.50 to 10.00; (0.977 µm)	0.3793
10.00 to 10.50; (0.691 µm)	0.3415
10.50 to 11.00; (0.488 µm)	0.2654
11.00 to 11.50; (0.345 µm)	0.0850
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000



**NMBAQCS - PS Exercise Data Workbook (Page 2 -  
Final Merged Data Submission)**

Return to APEM Ltd. by 06-02-15

Exercise Code:	PS55
LabCode:	PSA_2113_a
Sample Code:	PS552113

Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	0.0000
-4.00 to -3.50; 11.2 mm	1.7529
-3.50 to -3.00; 8 mm	49.9995
-3.00 to -2.50; 5.6 mm	25.9708
-2.50 to -2.00; 4 mm	2.3511
-2.00 to -1.50; 2.8 mm	0.3388
-1.50 to -1.00; 2 mm	0.0590
-1.00 to -0.50; 1.4 mm	0.0214
-0.50 to 0.00; 1 mm	0.0153
0.00 to 0.50; (707 µm)	0.0000
0.50 to 1.00; (500 µm)	0.0136
1.00 to 1.50; (353.6 µm)	0.2239
1.50 to 2.00; (250 µm)	0.7557
2.00 to 2.50; (176.8 µm)	0.7901
2.50 to 3.00; (125 µm)	0.3702
3.00 to 3.50; (88.39 µm)	0.0940
3.50 to 4.00; (62.5 µm)	0.1678
4.00 to 4.50; (44.19 µm)	0.5468
4.50 to 5.00; (31.25 µm)	0.9863
5.00 to 5.50; (22.097 µm)	1.3593
5.50 to 6.00; (15.625 µm)	1.7098
6.00 to 6.50; (11.049 µm)	2.0468
6.50 to 7.00; (7.813 µm)	2.2689
7.00 to 7.50; (5.524 µm)	2.2653
7.50 to 8.00; (3.906 µm)	1.9882
8.00 to 8.50; (2.762 µm)	1.4951
8.50 to 9.00; (1.953 µm)	0.9480
9.00 to 9.50; (1.381 µm)	0.5352
9.50 to 10.00; (0.977 µm)	0.3453
10.00 to 10.50; (0.691 µm)	0.2951
10.50 to 11.00; (0.488 µm)	0.2235
11.00 to 11.50; (0.345 µm)	0.0624
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000

**NMBAQCS - PS Exercise Data Workbook (Page 2 -  
Final Merged Data Submission)**

Return to APEM Ltd. by 06-02-15

Exercise Code:	<b>PS55</b>
LabCode:	<b>PSA_2114</b>
Sample Code:	<b>PS552114</b>

Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	0.0000
-4.00 to -3.50; 11.2 mm	1.8700
-3.50 to -3.00; 8 mm	41.5400
-3.00 to -2.50; 5.6 mm	32.6300
-2.50 to -2.00; 4 mm	3.5300
-2.00 to -1.50; 2.8 mm	0.3500
-1.50 to -1.00; 2 mm	0.0100
-1.00 to -0.50; 1.4 mm	0.0000
-0.50 to 0.00; 1 mm	0.0000
0.00 to 0.50; (707 µm)	0.0000
0.50 to 1.00; (500 µm)	0.1000
1.00 to 1.50; (353.6 µm)	0.4600
1.50 to 2.00; (250 µm)	0.7800
2.00 to 2.50; (176.8 µm)	0.7400
2.50 to 3.00; (125 µm)	0.5200
3.00 to 3.50; (88.39 µm)	0.4100
3.50 to 4.00; (62.5 µm)	0.4700
4.00 to 4.50; (44.19 µm)	0.6200
4.50 to 5.00; (31.25 µm)	0.8200
5.00 to 5.50; (22.097 µm)	1.1000
5.50 to 6.00; (15.625 µm)	1.3900
6.00 to 6.50; (11.049 µm)	1.7000
6.50 to 7.00; (7.813 µm)	2.1700
7.00 to 7.50; (5.524 µm)	2.3100
7.50 to 8.00; (3.906 µm)	2.0800
8.00 to 8.50; (2.762 µm)	1.7000
8.50 to 9.00; (1.953 µm)	1.0100
9.00 to 9.50; (1.381 µm)	0.5900
9.50 to 10.00; (0.977 µm)	0.4500
10.00 to 10.50; (0.691 µm)	0.6600
10.50 to 11.00; (0.488 µm)	0.0000
11.00 to 11.50; (0.345 µm)	0.0000
11.50 to 12.00; (0.244 µm)	0.0000
12.00 to 12.50; (0.173 µm)	0.0000
12.50 to 13.00; (0.122 µm)	0.0000
13.00 to 13.50; (0.086 µm)	0.0000

**NMBAQCS - PS Exercise Data Workbook (Page 2 -  
Final Merged Data Submission)**

Return to APEM Ltd. by 06-02-15

Exercise Code:	<b>PS55</b>
LabCode:	<b>PSA_BENCHMARK REPLICATE 1</b>
Sample Code:	<b>PS5521BMREP1</b>

Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	0.0000
-4.00 to -3.50; 11.2 mm	0.4590
-3.50 to -3.00; 8 mm	44.8010
-3.00 to -2.50; 5.6 mm	29.8093
-2.50 to -2.00; 4 mm	4.8226
-2.00 to -1.50; 2.8 mm	0.3501
-1.50 to -1.00; 2 mm	0.0011
-1.00 to -0.50; 1.4 mm	0.0000
-0.50 to 0.00; 1 mm	0.0000
0.00 to 0.50; (707 µm)	0.0000
0.50 to 1.00; (500 µm)	0.0000
1.00 to 1.50; (353.6 µm)	0.0577
1.50 to 2.00; (250 µm)	1.0225
2.00 to 2.50; (176.8 µm)	1.1812
2.50 to 3.00; (125 µm)	0.5664
3.00 to 3.50; (88.39 µm)	0.4177
3.50 to 4.00; (62.5 µm)	0.4099
4.00 to 4.50; (44.19 µm)	0.5743
4.50 to 5.00; (31.25 µm)	0.7050
5.00 to 5.50; (22.097 µm)	1.0447
5.50 to 6.00; (15.625 µm)	1.2158
6.00 to 6.50; (11.049 µm)	1.2802
6.50 to 7.00; (7.813 µm)	1.6170
7.00 to 7.50; (5.524 µm)	1.7727
7.50 to 8.00; (3.906 µm)	1.6244
8.00 to 8.50; (2.762 µm)	1.3001
8.50 to 9.00; (1.953 µm)	0.9580
9.00 to 9.50; (1.381 µm)	0.6042
9.50 to 10.00; (0.977 µm)	0.2975
10.00 to 10.50; (0.691 µm)	0.2492
10.50 to 11.00; (0.488 µm)	0.4546
11.00 to 11.50; (0.345 µm)	0.6842
11.50 to 12.00; (0.244 µm)	0.7090
12.00 to 12.50; (0.173 µm)	0.5207
12.50 to 13.00; (0.122 µm)	0.3259
13.00 to 13.50; (0.086 µm)	0.1640

**NMBAQCS - PS Exercise Data Workbook (Page 2 -  
Final Merged Data Submission)**

Return to APEM Ltd. by 06-02-15

Exercise Code:	<b>PS55</b>
LabCode:	<b>PSA_BENCHMARK REPLICATE 2</b>
Sample Code:	<b>PS5521BMREP2</b>

Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	0.0000
-4.00 to -3.50; 11.2 mm	0.6354
-3.50 to -3.00; 8 mm	43.0156
-3.00 to -2.50; 5.6 mm	30.8310
-2.50 to -2.00; 4 mm	4.8347
-2.00 to -1.50; 2.8 mm	0.2979
-1.50 to -1.00; 2 mm	0.0055
-1.00 to -0.50; 1.4 mm	0.0000
-0.50 to 0.00; 1 mm	0.0000
0.00 to 0.50; (707 µm)	0.0000
0.50 to 1.00; (500 µm)	0.0000
1.00 to 1.50; (353.6 µm)	0.0562
1.50 to 2.00; (250 µm)	1.0763
2.00 to 2.50; (176.8 µm)	1.2791
2.50 to 3.00; (125 µm)	0.4832
3.00 to 3.50; (88.39 µm)	0.3953
3.50 to 4.00; (62.5 µm)	0.3999
4.00 to 4.50; (44.19 µm)	0.5952
4.50 to 5.00; (31.25 µm)	0.7417
5.00 to 5.50; (22.097 µm)	1.0813
5.50 to 6.00; (15.625 µm)	1.2643
6.00 to 6.50; (11.049 µm)	1.3673
6.50 to 7.00; (7.813 µm)	1.7567
7.00 to 7.50; (5.524 µm)	1.9169
7.50 to 8.00; (3.906 µm)	1.7307
8.00 to 8.50; (2.762 µm)	1.3562
8.50 to 9.00; (1.953 µm)	0.9742
9.00 to 9.50; (1.381 µm)	0.6048
9.50 to 10.00; (0.977 µm)	0.3027
10.00 to 10.50; (0.691 µm)	0.2508
10.50 to 11.00; (0.488 µm)	0.4344
11.00 to 11.50; (0.345 µm)	0.6430
11.50 to 12.00; (0.244 µm)	0.6704
12.00 to 12.50; (0.173 µm)	0.5035
12.50 to 13.00; (0.122 µm)	0.3254
13.00 to 13.50; (0.086 µm)	0.1704

**NMBAQCS - PS Exercise Data Workbook (Page 2 -  
Final Merged Data Submission)**

Return to APEM Ltd. by 06-02-15

Exercise Code:	<b>PS55</b>
LabCode:	<b>PSA_BENCHMARK REPLICATE 3</b>
Sample Code:	<b>PS5521BMREP3</b>

Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	0.0000
-4.00 to -3.50; 11.2 mm	1.3880
-3.50 to -3.00; 8 mm	43.5484
-3.00 to -2.50; 5.6 mm	30.3366
-2.50 to -2.00; 4 mm	4.3146
-2.00 to -1.50; 2.8 mm	0.3865
-1.50 to -1.00; 2 mm	0.0011
-1.00 to -0.50; 1.4 mm	0.0000
-0.50 to 0.00; 1 mm	0.0000
0.00 to 0.50; (707 µm)	0.0000
0.50 to 1.00; (500 µm)	0.0000
1.00 to 1.50; (353.6 µm)	0.0388
1.50 to 2.00; (250 µm)	0.9282
2.00 to 2.50; (176.8 µm)	1.3412
2.50 to 3.00; (125 µm)	0.5715
3.00 to 3.50; (88.39 µm)	0.4167
3.50 to 4.00; (62.5 µm)	0.4101
4.00 to 4.50; (44.19 µm)	0.5650
4.50 to 5.00; (31.25 µm)	0.7094
5.00 to 5.50; (22.097 µm)	1.0364
5.50 to 6.00; (15.625 µm)	1.2178
6.00 to 6.50; (11.049 µm)	1.2882
6.50 to 7.00; (7.813 µm)	1.6769
7.00 to 7.50; (5.524 µm)	1.8467
7.50 to 8.00; (3.906 µm)	1.6621
8.00 to 8.50; (2.762 µm)	1.2914
8.50 to 9.00; (1.953 µm)	0.9364
9.00 to 9.50; (1.381 µm)	0.5847
9.50 to 10.00; (0.977 µm)	0.2709
10.00 to 10.50; (0.691 µm)	0.2330
10.50 to 11.00; (0.488 µm)	0.4777
11.00 to 11.50; (0.345 µm)	0.7437
11.50 to 12.00; (0.244 µm)	0.7644
12.00 to 12.50; (0.173 µm)	0.5419
12.50 to 13.00; (0.122 µm)	0.3213
13.00 to 13.50; (0.086 µm)	0.1501

**NMBAQCS - PS Exercise Data Workbook (Page 2 -  
Final Merged Data Submission)**

Return to APEM Ltd. by 06-02-15

Exercise Code:	<b>PS55</b>
LabCode:	<b>PSA_BENCHMARK REPLICATE 4</b>
Sample Code:	<b>PS5521BMREP4</b>

Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	0.0000
-4.00 to -3.50; 11.2 mm	1.1687
-3.50 to -3.00; 8 mm	42.7952
-3.00 to -2.50; 5.6 mm	31.0217
-2.50 to -2.00; 4 mm	4.6608
-2.00 to -1.50; 2.8 mm	0.3423
-1.50 to -1.00; 2 mm	0.0033
-1.00 to -0.50; 1.4 mm	0.0000
-0.50 to 0.00; 1 mm	0.0000
0.00 to 0.50; (707 µm)	0.0000
0.50 to 1.00; (500 µm)	0.0000
1.00 to 1.50; (353.6 µm)	0.0486
1.50 to 2.00; (250 µm)	0.9863
2.00 to 2.50; (176.8 µm)	1.2789
2.50 to 3.00; (125 µm)	0.6566
3.00 to 3.50; (88.39 µm)	0.4998
3.50 to 4.00; (62.5 µm)	0.4375
4.00 to 4.50; (44.19 µm)	0.5932
4.50 to 5.00; (31.25 µm)	0.7087
5.00 to 5.50; (22.097 µm)	1.0363
5.50 to 6.00; (15.625 µm)	1.2185
6.00 to 6.50; (11.049 µm)	1.3098
6.50 to 7.00; (7.813 µm)	1.6772
7.00 to 7.50; (5.524 µm)	1.8357
7.50 to 8.00; (3.906 µm)	1.6692
8.00 to 8.50; (2.762 µm)	1.3166
8.50 to 9.00; (1.953 µm)	0.9465
9.00 to 9.50; (1.381 µm)	0.5859
9.50 to 10.00; (0.977 µm)	0.2937
10.00 to 10.50; (0.691 µm)	0.2446
10.50 to 11.00; (0.488 µm)	0.4216
11.00 to 11.50; (0.345 µm)	0.6213
11.50 to 12.00; (0.244 µm)	0.6467
12.00 to 12.50; (0.173 µm)	0.4879
12.50 to 13.00; (0.122 µm)	0.3181
13.00 to 13.50; (0.086 µm)	0.1687

**NMBAQCS - PS Exercise Data Workbook (Page 2 -  
Final Merged Data Submission)**

Return to APEM Ltd. by 06-02-15

Exercise Code:	<b>PS55</b>
LabCode:	<b>PSA_BENCHMARK REPLICATE 5</b>
Sample Code:	<b>PS5521BMREP5</b>

Phi interval (explicit) + sieve mesh (theoretical sieves shown in brackets)	Volume/Weight (mark as "0" for not analysed or no material)
-6.50 to -6.00; 63 mm	0.0000
-6.00 to -5.50; 45 mm	0.0000
-5.50 to -5.00; 31.5 mm	0.0000
-5.00 to -4.50; 22.4 mm	0.0000
-4.50 to -4.00; 16 mm	0.0000
-4.00 to -3.50; 11.2 mm	0.5153
-3.50 to -3.00; 8 mm	43.0440
-3.00 to -2.50; 5.6 mm	32.2991
-2.50 to -2.00; 4 mm	4.0067
-2.00 to -1.50; 2.8 mm	0.3341
-1.50 to -1.00; 2 mm	0.0011
-1.00 to -0.50; 1.4 mm	0.0000
-0.50 to 0.00; 1 mm	0.0000
0.00 to 0.50; (707 µm)	0.0000
0.50 to 1.00; (500 µm)	0.0000
1.00 to 1.50; (353.6 µm)	0.0365
1.50 to 2.00; (250 µm)	0.9054
2.00 to 2.50; (176.8 µm)	1.3526
2.50 to 3.00; (125 µm)	0.6192
3.00 to 3.50; (88.39 µm)	0.3767
3.50 to 4.00; (62.5 µm)	0.3661
4.00 to 4.50; (44.19 µm)	0.5586
4.50 to 5.00; (31.25 µm)	0.7121
5.00 to 5.50; (22.097 µm)	1.0625
5.50 to 6.00; (15.625 µm)	1.2327
6.00 to 6.50; (11.049 µm)	1.2901
6.50 to 7.00; (7.813 µm)	1.6196
7.00 to 7.50; (5.524 µm)	1.7725
7.50 to 8.00; (3.906 µm)	1.6277
8.00 to 8.50; (2.762 µm)	1.3084
8.50 to 9.00; (1.953 µm)	0.9645
9.00 to 9.50; (1.381 µm)	0.6066
9.50 to 10.00; (0.977 µm)	0.3001
10.00 to 10.50; (0.691 µm)	0.2500
10.50 to 11.00; (0.488 µm)	0.4476
11.00 to 11.50; (0.345 µm)	0.6701
11.50 to 12.00; (0.244 µm)	0.6969
12.00 to 12.50; (0.173 µm)	0.5190
12.50 to 13.00; (0.122 µm)	0.3321
13.00 to 13.50; (0.086 µm)	0.1721

**Appendix 2. z-score calculations at each half phi-interval for participating laboratories and the benchmark average for PS55.**

	Phi interval														
	-6.50 to -6.00	-6.00 to -5.50	-5.50 to -5.00	-5.00 to -4.50	-4.50 to -4.00	-4.00 to -3.50	z-score	-3.50 to -3.00	z-score	-3.00 to -2.50	z-score	-2.50 to -2.00	z-score	-2.00 to -1.50	z-score
Benchmark Average	0.00	0.00	0.00	0.00	0.00	0.833	-0.23	43.441	0.26	30.860	-0.05	4.528	0.45	0.342	0.02
PSA_2101	0.00	0.00	0.00	0.00	0.00	0.218	-0.73	39.617	-0.74	33.184	0.75	6.190	1.88	0.399	1.00
PSA_2102	0.00	0.00	0.00	0.00	0.00	0.000	-0.91	42.946	0.13	29.921	-0.37	6.497	2.15	0.403	1.08
PSA_2103	0.00	0.00	0.00	0.00	0.00	0.433	-0.56	43.063	0.16	32.638	0.56	3.448	-0.48	0.325	-0.28
PSA_2105	0.00	0.00	0.00	0.00	0.00	0.459	-0.54	44.801	0.62	29.809	-0.40	4.823	0.70	0.350	0.16
PSA_2106	0.00	0.00	0.00	0.00	0.00	1.314	0.17	43.008	0.15	32.778	0.61	3.798	-0.18	0.433	1.59
PSA_2107	0.00	0.00	0.00	0.00	0.00	0.484	-0.52	41.282	-0.30	35.323	1.48	4.014	0.00	0.360	0.33
PSA_2108	0.00	0.00	0.00	0.00	0.00	0.000	-0.91	44.213	0.46	31.553	0.19	4.096	0.07	0.231	-1.91
PSA_2109	0.00	0.00	0.00	0.00	0.00	1.030	-0.07	41.935	-0.13	31.648	0.22	3.647	-0.31	0.253	-1.52
PSA_2110	0.00	0.00	0.00	0.00	0.00	1.210	0.08	42.260	-0.05	33.680	0.92	3.010	-0.86	0.330	-0.19
PSA_2111	0.00	0.00	0.00	0.00	0.00	0.000	-0.91	31.988	-2.73	25.643	-1.83	3.380	-0.54	0.352	0.18
PSA_2112	0.00	0.00	0.00	0.00	0.00	2.519	1.15	40.446	-0.52	32.722	0.59	4.245	0.20	0.393	0.90
PSA_2113	0.00	0.00	0.00	0.00	0.00	4.569	2.83	46.141	0.97	26.508	-1.53	2.586	-1.23	0.255	-1.49
PSA_2113_a	0.00	0.00	0.00	0.00	0.00	1.753	0.52	49.999	1.97	25.971	-1.72	2.351	-1.43	0.339	-0.04
PSA_2114	0.00	0.00	0.00	0.00	0.00	1.870	0.62	41.540	-0.24	32.630	0.56	3.530	-0.41	0.350	0.16
MEAN	0.00	0.00	0.00	0.00	0.00	1.1128		42.4453		30.9912		4.0094		0.3410	
STANDARD DEVIATION	0.00	0.00	0.00	0.00	0.00	1.2201		3.8280		2.9228		1.1598		0.0579	

All labs recorded zero therefore the mean and standard deviation were also zero.  
z < -1.96 or z > 1.96



**Appendix 2. z-score calculations at each half phi-interval for participating laboratories and the benchmark average for PS55.**

	Phi interval															
	-1.50 to -1.00	z-score	-1.00 to -0.50	z-score	-0.50 to 0.00	z-score	0.00 to 0.50	z-score	0.50 to 1.00	z-score	1.00 to 1.50	z-score	1.50 to 2.00	z-score	2.00 to 2.50	z-score
Benchmark Average	0.002	-0.45	0.000	-0.52	0.000	-0.53	0.000	-0.53	0.000	-0.63	0.048	-0.90	0.984	-0.09	1.287	0.73
PSA_2101	0.001	-0.54	0.001	-0.38	0.002	-0.03	0.009	-0.28	0.145	-0.24	0.587	-0.04	1.047	0.02	1.329	0.82
PSA_2102	0.023	0.91	0.020	2.20	0.008	1.30	0.028	0.21	0.176	-0.16	0.486	-0.21	0.770	-0.43	0.714	-0.58
PSA_2103	0.001	-0.55	0.000	-0.52	0.001	-0.29	0.000	-0.53	0.695	1.22	2.200	2.52	2.626	2.57	1.935	2.21
PSA_2105	0.001	-0.54	0.000	-0.52	0.000	-0.53	0.000	-0.53	0.000	-0.63	0.058	-0.89	1.023	-0.02	1.181	0.49
PSA_2106	0.010	0.06	0.000	-0.52	0.000	-0.46	0.134	3.07	1.353	2.98	1.719	1.75	0.760	-0.45	0.644	-0.74
PSA_2107	0.010	0.06	0.000	-0.52	0.000	-0.53	0.000	-0.53	0.002	-0.62	0.168	-0.71	0.340	-1.13	0.376	-1.36
PSA_2108	0.001	-0.55	0.001	-0.41	0.001	-0.29	0.021	0.05	0.388	0.41	0.945	0.52	1.208	0.28	1.007	0.09
PSA_2109	0.014	0.28	0.000	-0.52	0.000	-0.53	0.033	0.35	0.032	-0.55	0.612	0.00	2.150	1.80	1.484	1.18
PSA_2110	0.000	-0.61	0.010	0.85	0.000	-0.53	0.000	-0.53	0.110	-0.34	0.390	-0.36	0.580	-0.74	0.510	-1.05
PSA_2111	0.000	-0.61	0.000	-0.52	0.000	-0.53	0.070	1.35	0.514	0.74	1.038	0.67	1.320	0.46	1.043	0.17
PSA_2112	0.005	-0.27	0.001	-0.38	0.001	-0.28	0.000	-0.53	0.017	-0.59	0.189	-0.68	0.381	-1.06	0.373	-1.36
PSA_2113	0.002	-0.48	0.003	-0.11	0.005	0.67	0.000	-0.53	0.000	-0.63	0.101	-0.82	0.831	-0.33	1.114	0.33
PSA_2113_a	0.059	3.26	0.021	2.41	0.015	3.08	0.000	-0.53	0.014	-0.59	0.224	-0.62	0.756	-0.46	0.790	-0.41
PSA_2114	0.010	0.04	0.000	-0.52	0.000	-0.53	0.000	-0.53	0.100	-0.36	0.460	-0.25	0.780	-0.42	0.740	-0.52
MEAN	0.0093		0.0038		0.0022		0.0197		0.2364		0.6149		1.0370		0.9684	
STANDARD DEVIATION	0.0152		0.0073		0.0042		0.0373		0.3747		0.6292		0.6174		0.4372	

z < -1.96 or z > 1.96

**Appendix 2. z-score calculations at each half phi-interval for participating laboratories and the benchmark average for PS55.**

	Phi interval															
	2.50 to 3.00	z-score	3.00 to 3.50	z-score	3.50 to 4.00	z-score	4.00 to 4.50	z-score	4.50 to 5.00	z-score	5.00 to 5.50	z-score	5.50 to 6.00	z-score	6.00 to 6.50	z-score
Benchmark Average	0.579	0.13	0.421	0.43	0.405	-0.20	0.577	-0.40	0.72	-0.51	1.05	-0.23	1.23	-0.55	1.31	-0.45
PSA_2101	0.897	1.59	0.607	1.60	0.599	0.89	0.641	-0.33	0.80	-0.28	0.97	-0.42	1.21	-0.57	1.43	-0.28
PSA_2102	0.343	-0.96	0.135	-1.38	0.307	-0.75	0.666	-0.30	1.01	0.34	1.30	0.34	1.59	-0.03	1.90	0.41
PSA_2103	0.980	1.97	0.583	1.45	0.818	2.12	1.219	0.38	1.40	1.50	1.34	0.44	1.18	-0.63	1.01	-0.88
PSA_2105	0.566	0.07	0.418	0.40	0.410	-0.17	0.574	-0.41	0.70	-0.54	1.04	-0.25	1.22	-0.57	1.28	-0.49
PSA_2106	0.274	-1.28	0.452	0.62	0.465	0.14	0.817	-0.11	1.11	0.64	1.45	0.68	1.24	-0.53	1.47	-0.22
PSA_2107	0.320	-1.06	0.333	-0.13	0.446	0.03	0.642	-0.32	0.81	-0.25	1.03	-0.29	1.32	-0.42	1.71	0.13
PSA_2108	0.618	0.31	0.406	0.33	0.470	0.16	0.687	-0.27	0.92	0.08	1.12	-0.07	1.32	-0.42	1.53	-0.13
PSA_2109	0.733	0.84	0.150	-1.28	0.255	-1.04	3.750	3.49	0.00	-2.61	0.00	-2.67	3.80	3.10	0.00	-2.34
PSA_2110	0.380	-0.79	0.350	-0.02	0.460	0.11	0.640	-0.33	0.84	-0.15	1.13	-0.05	1.49	-0.18	1.83	0.30
PSA_2111	0.729	0.82	0.482	0.81	0.716	1.54	1.162	0.31	1.53	1.89	2.16	2.33	2.64	1.46	3.35	2.49
PSA_2112	0.338	-0.98	0.299	-0.34	0.446	0.03	0.579	-0.40	0.88	-0.03	1.11	-0.09	1.44	-0.25	1.84	0.32
PSA_2113	0.620	0.32	0.167	-1.18	0.174	-1.49	0.478	-0.53	0.84	-0.15	1.13	-0.06	1.46	-0.22	1.90	0.40
PSA_2113_a	0.370	-0.83	0.094	-1.64	0.168	-1.53	0.547	-0.44	0.99	0.28	1.36	0.48	1.71	0.13	2.05	0.62
PSA_2114	0.520	-0.14	0.410	0.35	0.470	0.17	0.620	-0.35	0.82	-0.21	1.10	-0.12	1.39	-0.32	1.70	0.12
MEAN	0.5512		0.3538		0.4405		0.9066		0.89		1.15		1.62		1.62	
STANDARD DEVIATION	0.2171		0.1586		0.1786		0.8147		0.34		0.43		0.70		0.69	

z < -1.96 or z > 1.96

**Appendix 2. z-score calculations at each half phi-interval for participating laboratories and the benchmark average for PS55.**

	Phi interval															
	6.50 to 7.00	z-score	7.00 to 7.50	z-score	7.50 to 8.00	z-score	8.00 to 8.50	z-score	8.50 to 9.00	z-score	9.00 to 9.50	z-score	9.50 to 10.00	z-score	10.00 to 10.50	z-score
Benchmark Average	1.67	-0.26	1.83	-0.17	1.66	-0.42	1.31	-0.22	0.96	-0.05	0.60	0.08	0.29	-0.25	0.25	-0.31
PSA_2101	1.63	-0.31	1.70	-0.31	1.54	-0.48	1.22	-0.36	0.88	-0.21	0.62	0.15	0.44	0.42	0.38	0.21
PSA_2102	2.19	0.36	2.36	0.37	2.27	-0.09	1.88	0.60	1.27	0.64	0.63	0.18	0.16	-0.84	0.00	-1.26
PSA_2103	0.88	-1.20	0.76	-1.27	0.65	-0.97	0.54	-1.34	0.41	-1.26	0.28	-0.99	0.21	-0.62	0.19	-0.52
PSA_2105	1.62	-0.33	1.77	-0.23	1.62	-0.44	1.30	-0.24	0.96	-0.05	0.60	0.11	0.30	-0.23	0.25	-0.29
PSA_2106	1.36	-0.63	0.98	-1.04	1.21	-0.66	1.06	-0.58	0.86	-0.26	0.49	-0.27	0.47	0.52	0.26	-0.24
PSA_2107	2.06	0.20	2.29	0.30	2.10	-0.18	1.60	0.19	1.09	0.23	0.67	0.35	0.42	0.29	0.38	0.22
PSA_2108	1.74	-0.18	1.84	-0.17	1.72	-0.39	1.37	-0.13	0.93	-0.10	0.58	0.01	0.40	0.21	0.34	0.06
PSA_2109	0.00	-2.24	0.00	-2.04	8.48	3.30	0.00	-2.11	0.00	-2.14	0.00	-1.94	0.00	-1.55	0.00	-1.26
PSA_2110	2.25	0.42	2.36	0.37	2.12	-0.17	1.62	0.22	0.99	0.03	0.51	-0.21	0.38	0.14	0.57	0.95
PSA_2111	3.91	2.39	4.31	2.36	3.99	0.86	3.10	2.36	2.18	2.64	1.45	2.97	0.97	2.76	0.96	2.46
PSA_2112	2.35	0.54	2.77	0.79	2.79	0.20	2.15	0.99	1.21	0.50	0.48	-0.32	0.03	-1.41	0.00	-1.26
PSA_2113	2.29	0.47	2.43	0.44	2.17	-0.14	1.61	0.21	0.99	0.03	0.56	-0.05	0.38	0.13	0.34	0.06
PSA_2113_a	2.27	0.44	2.27	0.27	1.99	-0.24	1.50	0.04	0.95	-0.07	0.54	-0.13	0.35	-0.02	0.30	-0.12
PSA_2114	2.17	0.33	2.31	0.32	2.08	-0.19	1.70	0.34	1.01	0.07	0.59	0.06	0.45	0.45	0.66	1.30
MEAN	1.89		2.00		2.43		1.46		0.98		0.57		0.35		0.33	
STANDARD DEVIATION	0.85		0.98		1.83		0.69		0.46		0.29		0.23		0.26	

z < -1.96 or z > 1.96

**Appendix 2. z-score calculations at each half phi-interval for participating laboratories and the benchmark average for PS55.**

	Phi interval											
	10.50 to 11.00	z-score	11.00 to 11.50	z-score	11.50 to 12.00	z-score	12.00 to 12.50	z-score	12.50 to 13.00	z-score	13.00 to 13.50	z-score
Benchmark Average	0.45	1.06	0.67	2.12	0.70	2.27	0.51	2.29	0.32	2.29	0.17	2.13
PSA_2101	0.39	0.80	0.40	1.00	0.36	0.95	0.26	0.92	0.17	0.95	0.12	1.49
PSA_2102	0.00	-1.00	0.00	-0.69	0.00	-0.46	0.00	-0.46	0.00	-0.46	0.00	-0.48
PSA_2103	0.15	-0.33	0.04	-0.50	0.00	-0.46	0.00	-0.46	0.00	-0.46	0.00	-0.48
PSA_2105	0.45	1.10	0.68	2.17	0.71	2.32	0.52	2.32	0.33	2.30	0.16	2.12
PSA_2106	0.07	-0.69	0.00	-0.69	0.00	-0.46	0.00	-0.46	0.00	-0.46	0.00	-0.48
PSA_2107	0.29	0.35	0.13	-0.16	0.00	-0.46	0.00	-0.46	0.00	-0.46	0.00	-0.48
PSA_2108	0.26	0.20	0.09	-0.31	0.00	-0.46	0.00	-0.46	0.00	-0.46	0.00	-0.48
PSA_2109	0.00	-1.00	0.00	-0.69	0.00	-0.46	0.00	-0.46	0.00	-0.46	0.00	-0.48
PSA_2110	0.00	-1.00	0.00	-0.69	0.00	-0.46	0.00	-0.46	0.00	-0.46	0.00	-0.48
PSA_2111	0.71	2.28	0.30	0.56	0.00	-0.46	0.00	-0.46	0.00	-0.46	0.00	-0.48
PSA_2112	0.00	-1.00	0.00	-0.69	0.00	-0.46	0.00	-0.46	0.00	-0.46	0.00	-0.48
PSA_2113	0.27	0.22	0.08	-0.33	0.00	-0.46	0.00	-0.46	0.00	-0.46	0.00	-0.48
PSA_2113_a	0.22	0.03	0.06	-0.43	0.00	-0.46	0.00	-0.46	0.00	-0.46	0.00	-0.48
PSA_2114	0.00	-1.00	0.00	-0.69	0.00	-0.46	0.00	-0.46	0.00	-0.46	0.00	-0.48
MEAN	0.22		0.16		0.12		0.09		0.05		0.03	
STANDARD DEVIATION	0.22		0.24		0.26		0.19		0.12		0.06	

z < -1.96 or z > 1.96